



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB110430

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC OH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.4	4.01	08/07/2020	17:15
2	CAL2	1	Water	NA	NA	21.4	7.00	08/07/2020	17:16
3	CAL3	1	Water	NA	NA	21.4	10.03	08/07/2020	17:19
4	ICV	1	Water	NA	NA	21.4	6.98	08/07/2020	17:20
5	CCV1	1	Water	NA	NA	21.4	2.01	08/07/2020	17:22
6	L3506-05	1	Solid	20.03	20	26.5	8.57	08/07/2020	17:25
7	L3506-05DUP	1	Solid	20.04	20	26.6	8.59	08/07/2020	17:26
8	CCV2	1	Water	NA	NA	21.3	12.02	08/07/2020	17:30

WORKLIST(Hardcopy Internal Chain)

VB 110430

WorkList Name : corrosivity I3506

WorkList ID : 141274

Department : Wet-Chemistry

Date : 08-07-2020 15:39:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/07/2020	Solid	L3506-05	Corrosivity	Cool 4 deg C	PSEG05	J22	OR-02-080720	08/07/2020	9045D

Date/Time 08/07/2020 15:45

Received by: [Signature]

Relinquished by: KS. (wetchem)

Date/Time

Received by:

Relinquished by:

08/07/2020 17:30
KS (wetchem)
78 (wetchem)