


L3110443

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.069	-0.069	-14	0.000	-14				
0.2	0.200	0.211	0.011	787	0.0 %	787				
2.5	2.500	2.466	-0.034	7234	0.0 %	7234				
5.0	5.000	5.050	0.050	14624	0.0 %	14624				
7.5	7.500	7.801	0.301	22493	0.0 %	22493				
10.0	10.000	9.741	-0.259	28042	0.0 %	28042				

LB110443 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Aug 2020 10:05:47

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-14	-	0.0000	7470A	PPB	10 Aug 2020 10:51:31	S	Std
0.2 - 1	50.2	787	-	0.2000	7470A	PPB	10 Aug 2020 10:53:38	S	Std
2.5 - 1	52.5	7234	-	2.5000	7470A	PPB	10 Aug 2020 10:55:46	S	Std
5.0 - 1	55	14624	-	5.0000	7470A	PPB	10 Aug 2020 10:57:53	S	Std
7.5 - 1	57.5	22493	-	7.5000	7470A	PPB	10 Aug 2020 11:00:01	S	Std
10.0 - 1	59.0	28042	-	10.0000	7470A	PPB	10 Aug 2020 11:09:46	S	Std
ICV61 - 1	ICV61	11833	4.0737	-7470A	PPB	PPB	10 Aug 2020 11:12:25	U	SMPL
ICB61 - 1	ICB61	71	-0.0390	-7470A	PPB	PPB	10 Aug 2020 11:16:52	U	SMPL
CCV79 - 1	CCV79	15867	5.4842	-7470A	PPB	PPB	10 Aug 2020 11:18:57	U	SMPL
CCB79 - 1	CCB79	83	-0.0348	-7470A	PPB	PPB	10 Aug 2020 11:23:31	U	SMPL
CRA - 1	CRA	667	0.1694	-7470A	PPB	PPB	10 Aug 2020 11:25:36	U	SMPL
HighStd - 1	HighStd	28343	9.8465	-7470A	PPB	PPB	10 Aug 2020 11:27:42	U	SMPL
ChkStd - 1	ChkStd	22157	7.6835	-7470A	PPB	PPB	10 Aug 2020 11:29:48	U	SMPL
PB130823BL - 1	PBW	95	-0.0306	-7470A	PPB	PPB	10 Aug 2020 11:35:50	U	SMPL
PB130823BS - 1	LCSW	11705	4.0289	-7470A	PPB	PPB	10 Aug 2020 11:37:56	U	SMPL
L3362-07 - 1	RBR-59	-82	-0.0925	-7470A	PPB	PPB	10 Aug 2020 11:40:03	U	SMPL
L3508-04 - 1	TR-04-080620	-109	-0.1019	-7470A	PPB	PPB	10 Aug 2020 11:42:11	U	SMPL
L3508-04DUP - 1	TR-04-080620DUP	-27	-0.0733	-7470A	PPB	PPB	10 Aug 2020 11:44:18	U	SMPL
L3508-04MS - 1	TR-04-080620MS	13642	4.7062	-7470A	PPB	PPB	10 Aug 2020 11:46:26	U	SMPL
L3508-04MSD - 1	TR-04-080620MSD	12926	4.4558	-7470A	PPB	PPB	10 Aug 2020 11:48:33	U	SMPL
CCV80 - 1	CCV80	15877	5.4877	-7470A	PPB	PPB	10 Aug 2020 11:50:40	U	SMPL
CCB80 - 1	CCB80	95	-0.0306	-7470A	PPB	PPB	10 Aug 2020 11:55:20	U	SMPL
L3566-01 - 1	31157-60-94	-26	-0.0729	-7470A	PPB	PPB	10 Aug 2020 11:57:25	U	SMPL
L3594-02 - 1	TP-1	-27	-0.0733	-7470A	PPB	PPB	10 Aug 2020 11:59:31	U	SMPL
L3594-06 - 1	TP-2	-30	-0.0743	-7470A	PPB	PPB	10 Aug 2020 12:01:36	U	SMPL
L3594-10 - 1	TP-3	-25	-0.0726	-7470A	PPB	PPB	10 Aug 2020 12:03:43	U	SMPL
PB130823TB - 1	PB130789TB	-16	-0.0694	-7470A	PPB	PPB	10 Aug 2020 12:05:49	U	SMPL
L3508-04LX5 - 1		-8	-0.0666	-7470A	PPB	PPB	10 Aug 2020 12:07:56	U	SMPL
L3508-04A - 1		12764	4.3992	-7470A	PPB	PPB	10 Aug 2020 12:10:03	U	SMPL
CCV81 - 1	CCV81	15273	5.2765	-7470A	PPB	PPB	10 Aug 2020 12:14:39	U	SMPL
CCB81 - 1	CCB81	45	-0.0481	-7470A	PPB	PPB	10 Aug 2020 12:19:04	U	SMPL