



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110451

Slope : 98.8

pH Meter ID : WC OH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.4	4.01	08/10/2020	14:40
2	CAL2	1	Water	NA	NA	21.4	7.01	08/10/2020	14:41
3	CAL3	1	Water	NA	NA	21.4	10.03	08/10/2020	14:42
4	ICV	1	Water	NA	NA	21.4	6.98	08/10/2020	14:43
5	CCV1	1	Water	NA	NA	21.4	2.02	08/10/2020	14:45
6	L3626-01	1	Water	NA	NA	27.2	7.60	08/10/2020	14:48
7	L3626-01DUP	1	Water	NA	NA	27.3	7.61	08/10/2020	14:49
8	CCV2	1	Water	NA	NA	21.5	12.02	08/10/2020	14:55

LB 110451

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3626 WorkList ID : 141326 Department : Wet-Chemistry Date : 08-10-2020 14:32:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/10/2020	Water	L3626-01	pH	Cool 4 deg C	PSEG01	J33	1717	08/10/2020	9040C

Date/Time 08.10.2020 14.35
 Received by: [Signature]
 Relinquished by: [Signature]

Date/Time 08.10.2020 15.35
 Received by: [Signature]
 Relinquished by: [Signature]