



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : ketankumar

Run Number: LB110477

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC OH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.6	4.00	08/11/2020	17:10
2	CAL2	1	Water	NA	NA	21.6	7.01	08/11/2020	17:11
3	CAL3	1	Water	NA	NA	21.6	10.03	08/11/2020	17:12
4	ICV	1	Water	NA	NA	21.6	6.98	08/11/2020	17:15
5	CCV1	1	Water	NA	NA	21.6	2.02	08/11/2020	17:16
6	L3504-06	1	Solid	20.03	20	27.2	9.11	08/11/2020	17:19
7	L3504-06DUP	1	Solid	20.04	20	27.2	9.12	08/11/2020	17:20
8	L3508-10	1	Solid	20.03	20	27.3	7.73	08/11/2020	17:22
9	L3637-08	1	Solid	20.02	20	26.6	7.60	08/11/2020	17:23
10	L3637-14	1	Solid	20.03	20	26.4	8.41	08/11/2020	17:25
11	L3637-20	1	Solid	20.04	20	26.1	8.25	08/11/2020	17:30
12	L3645-01	1	Solid	20.02	20	26.1	9.92	08/11/2020	17:32
13	CCV2	1	Water	NA	NA	21.3	12.02	08/11/2020	17:35

WORKLIST(Hardcopy Internal Chain)

103110477

WorkList Name : corrosivity I3608

WorkList ID : 141376

Department : Wet-Chemistry

Date : 08-11-2020 15:02:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/11/2020	Solid	L3504-06	Corrosivity	Cool 4 deg C	PSEG05	J52	EO-02-081120	08/11/2020	9045D
08/11/2020	Solid	L3508-10	Corrosivity	Cool 4 deg C	PSEG05		TR-05-081120	08/11/2020	9045D
08/10/2020	Solid	L3637-08	Corrosivity	Cool 4 deg C	TULL02	J33	CAR-WASH-AREA-1	08/10/2020	9045D
08/10/2020	Solid	L3637-14	Corrosivity	Cool 4 deg C	TULL02	J33	CAR-WASH-AREA-2	08/10/2020	9045D
08/10/2020	Solid	L3637-20	Corrosivity	Cool 4 deg C	TULL02	J33	CAR-WASH-AREA-3	08/10/2020	9045D
08/10/2020	Solid	L3645-01	Corrosivity	Cool 4 deg C	DIST01	J51	WC-3	08/10/2020	9045D

Date/Time 08/11/2020 16:00

Received by: Jo (wet-chem)

Relinquished by: JS (wet-chem)

Date/Time 08/11/2020 18:00

Received by: JS (wet-chem)

Relinquished by: Jo (wet-chem)