

LB110492


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

32693



A= 0.0000e+000

B= 3.0137e-004

C= 4.0650e-002

Rho= 0.9995592

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.054	0.054	45	0.000	45				
0.2	0.200	0.195	-0.005	511	0.0 %	511				
2.5	2.500	2.431	-0.069	7930	0.0 %	7930				
5.0	5.000	4.914	-0.086	16171	0.0 %	16171				
7.5	7.500	7.713	0.213	25459	0.0 %	25459				
10.0	10.000	9.893	-0.107	32693	0.0 %	32693				

LB110492 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Aug 2020 12:51:51

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	45	-	0.0000	7470A	PPB	12 Aug 2020 13:26:40	S	Std
0.2 - 1	50.2	511	-	0.2000	7470A	PPB	12 Aug 2020 13:28:48	S	Std
2.5 - 1	52.5	7930	-	2.5000	7470A	PPB	12 Aug 2020 13:30:55	S	Std
5.0 - 1	55	16171	-	5.0000	7470A	PPB	12 Aug 2020 13:33:03	S	Std
7.5 - 1	57.5	25459	-	7.5000	7470A	PPB	12 Aug 2020 13:40:00	S	Std
10.0 - 1	510	32693	-	10.0000	7470A	PPB	12 Aug 2020 13:42:07	S	Std
ICV65 - 1	ICV65	12957	3.9455	-7470A	PPB	PPB	12 Aug 2020 13:47:04	U	SMPL
ICB65 - 1	ICB65	-237	-0.0308	-7470A	PPB	PPB	12 Aug 2020 13:49:10	U	SMPL
CCV04 - 1	CCV04	14987	4.5573	-7470A	PPB	PPB	12 Aug 2020 13:51:16	U	SMPL
CCB04 - 1	CCB04	-210	-0.0226	-7470A	PPB	PPB	12 Aug 2020 13:53:22	U	SMPL
CRA - 1	CRA	434	0.1714	-7470A	PPB	PPB	12 Aug 2020 13:55:29	U	SMPL
HighStd - 1	HighStd	29714	8.9955	-7470A	PPB	PPB	12 Aug 2020 13:57:35	U	SMPL
ChkStd - 1	ChkStd	21465	6.5095	-7470A	PPB	PPB	12 Aug 2020 13:59:41	U	SMPL
PB130899BL - 1	PBW	-356	-0.0666	-7470A	PPB	PPB	12 Aug 2020 14:01:49	U	SMPL
PB130899BS - 1	LCSW	12252	3.7330	-7470A	PPB	PPB	12 Aug 2020 14:06:20	U	SMPL
L3624-03 - 1	SS-TP-1	-239	-0.0314	-7470A	PPB	PPB	12 Aug 2020 14:08:27	U	SMPL
L3624-03DUP - 1	SS-TP-1DUP	-59	0.0229	-7470A	PPB	PPB	12 Aug 2020 14:10:35	U	SMPL
L3624-03MS - 1	SS-TP-1MS	11414	3.4805	-7470A	PPB	PPB	12 Aug 2020 14:12:43	U	SMPL
L3624-03MSD - 1	SS-TP-1MSD	11749	3.5814	-7470A	PPB	PPB	12 Aug 2020 14:14:51	U	SMPL
L3624-08 - 1	SS-TP-2	-213	-0.0235	-7470A	PPB	PPB	12 Aug 2020 14:16:57	U	SMPL
CCV05 - 1	CCV05	16276	4.9457	-7470A	PPB	PPB	12 Aug 2020 14:19:04	U	SMPL
CCB05 - 1	CCB05	-199	-0.0193	-7470A	PPB	PPB	12 Aug 2020 14:21:10	U	SMPL
L3624-13 - 1	SS-TP-3	-54	0.0244	-7470A	PPB	PPB	12 Aug 2020 14:23:17	U	SMPL
L3624-18 - 1	SS-TP-4	-12	0.0370	-7470A	PPB	PPB	12 Aug 2020 14:25:23	U	SMPL
PB130899TB - 1	PB130873TB	-24	0.0334	-7470A	PPB	PPB	12 Aug 2020 14:27:29	U	SMPL
L3624-03LX5 - 1		-11	0.0373	-7470A	PPB	PPB	12 Aug 2020 14:29:35	U	SMPL
L3624-03A - 1		11930	3.6360	-7470A	PPB	PPB	12 Aug 2020 14:31:42	U	SMPL
L3620-01 - 1	5-GA-BUCKET	41996	12.6970	-7470A	PPB	PPB	12 Aug 2020 14:33:49	U	SMPL
L3620-01DLX2 - 1		19935	6.0484	-7470A	PPB	PPB	12 Aug 2020 14:35:57	U	SMPL
CCV06 - 1	CCV06	14919	4.5368	-7470A	PPB	PPB	12 Aug 2020 14:38:07	U	SMPL
CCB06 - 1	CCB06	-259	-0.0374	-7470A	PPB	PPB	12 Aug 2020 14:40:17	U	SMPL