


LB116500

ISM02.4

INSTRUMENT ID: CY1



Linear

A= 0.0000e+000

B= 2.6239e-004

C= 1.1095e-002

Rho= 0.9999726

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.033	0.033	83	0.000	83					1
0.05	0.050					159					6
0.20	0.200	0.212	0.012	767	0.0 %	767					2
2.50	2.500	2.450	-0.050	9297	0.0 %	9297					0
5.00	5.000	4.979	-0.021	18932	0.0 %	18932					0
7.50	7.500	7.513	0.013	28592	0.0 %	28592					0
10.0	10.000	10.012	0.012	38117	0.0 %	38117					0

LB110500

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	83	-		0 ISM02.4	PPB	8/12/2020 15:52	Std
	0.2 S0.2	767	-		0.2 ISM02.4	PPB	8/12/2020 15:54	Std
	2.5 S2.5	9297	-		2.5 ISM02.4	PPB	8/12/2020 15:57	Std
	5 S5.0	18932	-		5 ISM02.4	PPB	8/12/2020 15:59	Std
	7.5 S7.5	28592	-		7.5 ISM02.4	PPB	8/12/2020 16:01	Std
	10 S10.0	38117	-		10 ISM02.4	PPB	8/12/2020 16:06	Std
ICV	ICV	14234	3.7459 -		ISM02.4	PPB	8/12/2020 16:09	SMPL
ICB	ICB	-143	-0.0264 -		ISM02.4	PPB	8/12/2020 16:11	SMPL
CCV007	CCV007	19446	5.1135 -		ISM02.4	PPB	8/12/2020 16:13	SMPL
CCB007	CCB007	-149	-0.028 -		ISM02.4	PPB	8/12/2020 16:16	SMPL
PB130927BL	PBW927	9	0.0135 -		ISM02.4	PPB	8/12/2020 16:18	SMPL
L3646-01	MCB7H5	1306	0.3538 -		ISM02.4	PPB	8/12/2020 16:20	SMPL
L3646-02	MCB7H6	316	0.094 -		ISM02.4	PPB	8/12/2020 16:22	SMPL
L3646-03	MCB7H7	132	0.0457 -		ISM02.4	PPB	8/12/2020 16:25	SMPL
L3646-04	MCB7H7D	171	0.056 -		ISM02.4	PPB	8/12/2020 16:27	SMPL
L3646-05	MCB7H7S	4066	1.078 -		ISM02.4	PPB	8/12/2020 16:29	SMPL
CCV008	CCV008	18458	4.8542 -		ISM02.4	PPB	8/12/2020 16:31	SMPL
CCB008	CCB008	-190	-0.0388 -		ISM02.4	PPB	8/12/2020 16:34	SMPL