



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110505

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC OH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	08/12/2020	17:37
2	CAL2	1	Water	NA	NA	21.3	7.00	08/12/2020	17:38
3	CAL3	1	Water	NA	NA	21.3	10.03	08/12/2020	17:39
4	ICV	1	Water	NA	NA	21.3	6.98	08/12/2020	17:40
5	CCV1	1	Water	NA	NA	21.3	2.02	08/12/2020	17:42
6	L3663-03	1	Solid	20.03	20	27.2	5.60	08/12/2020	18:05
7	L3663-04	1	Solid	20.02	20	26.7	6.92	08/12/2020	18:06
8	L3663-05	1	Solid	20.03	20	27.2	5.80	08/12/2020	18:08
9	L3663-06	1	Solid	20.02	20	27.2	6.28	08/12/2020	18:10
10	L3663-07	1	Solid	20.04	20	27.3	6.30	08/12/2020	18:11
11	L3663-08	1	Solid	20.02	20	27.3	6.25	08/12/2020	18:12
12	L3678-02	1	Solid	20.03	20	27.8	8.51	08/12/2020	18:15
13	L3678-02DUP	1	Solid	20.04	20	27.9	8.52	08/12/2020	18:16
14	CCV2	1	Water	NA	NA	21.3	12.02	08/12/2020	18:19

✓ 110505

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3663 WorkList ID : 141415 Department : Wet-Chemistry Date : 08-12-2020 17:44:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/12/2020	Solid	L3663-03	pH	Cool 4 deg C	PSEG01	K12	31001	08/12/2020	9045D
08/12/2020	Solid	L3663-04	pH	Cool 4 deg C	PSEG01	K12	31022	08/12/2020	9045D
08/12/2020	Solid	L3663-05	pH	Cool 4 deg C	PSEG01	K12	31023	08/12/2020	9045D
08/12/2020	Solid	L3663-06	pH	Cool 4 deg C	PSEG01	K12	31024	08/12/2020	9045D
08/12/2020	Solid	L3663-07	pH	Cool 4 deg C	PSEG01	K12	31025	08/12/2020	9045D
08/12/2020	Solid	L3663-08	pH	Cool 4 deg C	PSEG01	K12	31026	08/12/2020	9045D
08/12/2020	Solid	L3678-02	pH	Cool 4 deg C	JPCL01	J61	BROOKLYN-BRIDGE	08/12/2020	9045D

Date/Time 08.12.2020 16:30
 Received by: JP (Wet) Unlmy
 Relinquished by: KS (Wetchem)

Date/Time 08.12.2020 16:30
 Received by: KS (Wetchem)
 Relinquished by: JP (Wet) Chlmy