


LB110316

7471B

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.1564e-004

C= 1.1499e-003

Rho= 0.9999112

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.012	0.012	48	0.000	48				
0.2	0.200	0.195	-0.005	901	0.0 %	901				
2.5	2.500	2.431	-0.069	11268	0.0 %	11268				
5.0	5.000	5.071	0.071	23510	0.0 %	23510				
7.5	7.500	7.543	0.043	34976	0.0 %	34976				
10.0	10.000	9.948	-0.052	46128	0.0 %	46128				

LB110516 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 13 Aug 2020 13:47:10

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	48	-	0.00007471B	PPB	13 Aug 2020 13:49:35	S	Std
0.2 - 1	50.2	901	-	0.20007471B	PPB	13 Aug 2020 13:51:52	S	Std
2.5 - 1	52.5	11268	-	2.50007471B	PPB	13 Aug 2020 13:54:09	S	Std
5.0 - 1	55	23510	-	5.00007471B	PPB	13 Aug 2020 13:56:26	S	Std
7.5 - 1	57.5	34976	-	7.50007471B	PPB	13 Aug 2020 14:01:17	S	Std
10.0 - 1	57.8	46128	-	10.00007471B	PPB	13 Aug 2020 14:06:08	S	Std
ICV67 - 1	ICV67	18557	4.0027	-7471B	PPB	13 Aug 2020 14:08:49	U	SMPL
ICB67 - 1	ICB67	-173	-0.0362	-7471B	PPB	13 Aug 2020 14:11:05	U	SMPL
CCV14 - 1	CCV14	24306	5.2424	-7471B	PPB	13 Aug 2020 14:13:21	U	SMPL
CCB14 - 1	CCB14	-190	-0.0398	-7471B	PPB	13 Aug 2020 14:15:36	U	SMPL
CRA - 1	CRA	877	0.1903	-7471B	PPB	13 Aug 2020 14:17:53	U	SMPL
HighStd - 1	HighStd	45872	9.8929	-7471B	PPB	13 Aug 2020 14:20:07	U	SMPL
ChkStd - 1	ChkStd	34356	7.4096	-7471B	PPB	13 Aug 2020 14:22:22	U	SMPL
PB130939BL - 1	PBS	-261	-0.0551	-7471B	PPB	13 Aug 2020 14:24:41	U	SMPL
PB130939BS - 1	LCSS	18943	4.0860	-7471B	PPB	13 Aug 2020 14:26:59	U	SMPL
L3501-01 - 1	NB-08-081220	1404	0.3039	-7471B	PPB	13 Aug 2020 14:29:15	U	SMPL
L3508-13 - 1	TR-04-081220	1081	0.2343	-7471B	PPB	13 Aug 2020 14:31:33	U	SMPL
L3508-13DUP - 1	TR-04-081220DUP	943	0.2045	-7471B	PPB	13 Aug 2020 14:33:51	U	SMPL
L3508-13MS - 1	TR-04-081220MS	20202	4.3575	-7471B	PPB	13 Aug 2020 14:36:09	U	SMPL
L3508-13MSD - 1	TR-04-081220MSD	20048	4.3242	-7471B	PPB	13 Aug 2020 14:38:24	U	SMPL
CCV15 - 1	CCV15	23998	5.1760	-7471B	PPB	13 Aug 2020 14:40:41	U	SMPL
CCB15 - 1	CCB15	-311	-0.0659	-7471B	PPB	13 Aug 2020 14:42:59	U	SMPL
L3658-01 - 1	VNJ-235	1182	0.2560	-7471B	PPB	13 Aug 2020 14:45:15	U	SMPL
L3659-01 - 1	SP1-1	4392	0.9482	-7471B	PPB	13 Aug 2020 14:47:30	U	SMPL
L3659-04 - 1	SP1-2	5026	1.0849	-7471B	PPB	13 Aug 2020 14:49:46	U	SMPL
L3659-07 - 1	SP1-3	1278	0.2767	-7471B	PPB	13 Aug 2020 14:52:01	U	SMPL
L3661-02 - 1	RBR-53	5901	1.2736	-7471B	PPB	13 Aug 2020 14:54:17	U	SMPL
L3677-02 - 1	TP-6	557918	HIGH	-7471B	PPB	13 Aug 2020 14:56:34	X	
L3677-05 - 1	TP-1	176511	38.0635	-7471B	PPB	13 Aug 2020 14:58:26	U	SMPL
L3508-13LX5 - 1		56	0.0132	-7471B	PPB	13 Aug 2020 15:03:52	U	SMPL
L3508-13A - 1		20329	4.3848	-7471B	PPB	13 Aug 2020 15:06:08	U	SMPL
L3677-02DLX10 - 1		93388	20.1391	-7471B	PPB	13 Aug 2020 15:08:24	U	SMPL
CCV16 - 1	CCV16	23826	5.1389	-7471B	PPB	13 Aug 2020 15:10:42	U	SMPL
CCB16 - 1	CCB16	-380	-0.0808	-7471B	PPB	13 Aug 2020 15:13:01	U	SMPL
L3677-02DL2X50 - 1		18427	3.9747	-7471B	PPB	13 Aug 2020 15:15:19	U	SMPL
L3677-05DLX10 - 1		17999	3.8824	-7471B	PPB	13 Aug 2020 15:17:35	U	SMPL
CCV17 - 1	CCV17	23816	5.1368	-7471B	PPB	13 Aug 2020 15:19:52	U	SMPL
CCB17 - 1	CCB17	-291	-0.0616	-7471B	PPB	13 Aug 2020 15:22:09	U	SMPL