



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB110522

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC OH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.4	4.01	08/13/2020	16:40
2	CAL2	1	Water	NA	NA	21.4	7.01	08/13/2020	16:41
3	CAL3	1	Water	NA	NA	21.4	10.04	08/13/2020	16:43
4	ICV	1	Water	NA	NA	21.4	6.99	08/13/2020	16:45
5	CCV1	1	Water	NA	NA	21.4	2.01	08/13/2020	16:46
6	L3502-10	1	Solid	20.03	20	24.6	8.81	08/13/2020	16:48
7	L3682-06	1	Solid	20.02	20	25.0	11.24	08/13/2020	16:50
8	L3682-07	1	Solid	20.03	20	24.8	11.10	08/13/2020	16:51
9	L3682-08	1	Solid	20.04	20	24.5	10.96	08/13/2020	17:00
10	L3682-09	1	Solid	20.04	20	24.4	11.03	08/13/2020	17:02
11	L3682-10	1	Solid	20.03	20	24.3	11.26	08/13/2020	17:05
12	L3682-10DUP	1	Solid	20.04	20	24.4	11.27	08/13/2020	17:06
13	CCV2	1	Water	NA	NA	21.4	12.03	08/13/2020	17:10

WORKLIST(Hardcopy Internal Chain)

18 110522

WorkList Name : corrosivity I3682 WorkList ID : 141442 Department : Wet-Chemistry Date : 08-13-2020 14:19:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/13/2020	Solid	L3502-10	Corrosivity	Cool 4 deg C	PSEG05		OK-01-081320	08/13/2020	9045D
08/12/2020	Solid	L3682-06	Corrosivity	Cool 4 deg C	AQUA04	J61	20200513-2801-A	08/12/2020	9045D
08/12/2020	Solid	L3682-07	Corrosivity	Cool 4 deg C	AQUA04	J61	20200514-2801-B	08/12/2020	9045D
08/12/2020	Solid	L3682-08	Corrosivity	Cool 4 deg C	AQUA04	J61	20200515-2801-C	08/12/2020	9045D
08/12/2020	Solid	L3682-09	Corrosivity	Cool 4 deg C	AQUA04	J61	20200516-2801-D	08/12/2020	9045D
08/12/2020	Solid	L3682-10	Corrosivity	Cool 4 deg C	AQUA04	J61	20200517-2801-E	08/12/2020	9045D

Date/Time 08-13-2020 15:40
 Received by: Jb (Wetchem)
 Relinquished by: RC (Ext Lab)

Date/Time 08-13-2020 17:20
 Received by: RC (Ext Lab)
 Relinquished by: Jb (Wetchem)