


Batch Folder: D:\Agilent\ICPMH\1\DATA\P8081320-MS.b\
Analysis File: P8081320-MS.batch.bin
DA Date-Time: 2020-08-13 15:57:30
Calibration Title:
Calibration Method: External Calibration
VIS Interpolation Fit:

Level	Standard Data File	Sample Name	Acq. Date-Time
1	003CALB.d	S00	2020-08-13 15:32:12
2	005CALS.d	S02	2020-08-13 15:38:11
3	006CALS.d	S03	2020-08-13 15:41:11
4	007CALS.d	S04	2020-08-13 15:44:09
5	008CALS.d	S05	2020-08-13 15:47:10
6	009CALS.d	S06	2020-08-13 15:50:08
7	010CALS.d	S07	2020-08-13 15:53:09
8	011CALS.d	S08	2020-08-13 15:56:07

9 Be [No Gas] ISTD:6 Li [No Gas]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	29.99	0.0000	P	25.5
2	☐	1.000	0.971	889.86	0.0007	P	4.6
3	☐	50.000	51.292	47307.20	0.0378	P	2.6
4	☐	125.000	127.811	120323.83	0.0940	P	1.4
5	☐	250.000	250.911	240849.56	0.1846	P	1.1
6	☐	500.000	494.203	472903.86	0.3635	P	1.4
7	☐	1000.000	1002.255	930943.91	0.7372	P	1.4
8	☐			293.95	0.0002	P	3.8

$$y = 7.3550\text{E-}004 * x + 2.4537\text{E-}005$$


$$R = 1.0000$$

$$DL = 0.02557$$

Weight: <None>

Min Conc: 0

10 B [No Gas] ISTD:6 Li [No Gas]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	125.98	0.0001	P	15.2
2	☐	25.000	20.697	1764.43	0.0015	P	10.0
3	☐	50.000	47.505	4039.78	0.0032	P	11.3
4	☐	125.000	120.726	10282.92	0.0080	P	7.7
5	☐	250.000	240.036	20694.42	0.0159	P	5.7
6	☐	500.000	486.562	41657.22	0.0320	P	9.1
7	☐	1000.000	1009.977	83836.72	0.0664	P	8.2
8	☐			7695.85	0.0055	P	4.9


$$y = 6.5625\text{E-}005 * x + 1.0354\text{E-}004$$

$$R = 0.9998$$

$$DL = 0.7181$$

Weight: <None>

Min Conc: 0

11 B [No Gas] ISTD:6 Li [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	591.23	0.0005	P	8.1
2	☐	25.000	21.592	8897.64	0.0074	P	9.5
3	☐	50.000	48.186	19892.87	0.0159	P	8.6
4	☐	125.000	121.385	50244.39	0.0392	P	8.3
5	☐	250.000	242.202	101494.51	0.0778	P	7.6
6	☐	500.000	485.614	202114.86	0.1554	P	8.5
7	☐	1000.000	1009.770	407473.86	0.3227	P	7.4
8	☐			36277.61	0.0259	P	3.5


$$y = 3.1905E-004 * x + 4.8727E-004$$

$$R = 0.9998$$

$$DL = 0.369$$

Weight: <None>

Min Conc: 0

23 Na [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	54304.14	0.4052	P	1.0
2	☐	500.000	497.091	129386.85	0.9629	P	1.9
3	☐	5000.000	5748.738	887407.59	6.8554	P	7.4
4	☐	12500.000	13183.104	2103285.68	15.1970	A	1.3
5	☐	25000.000	26053.301	3980350.87	29.6377	A	1.0
6	☐	50000.000	53118.186	7919269.88	60.0053	A	1.1
7	☐	100000.00	103157.52	15844232.13	116.1509	A	0.8
8	☐	500000.00	498979.45	78033651.74	560.2742	A	1.4


$$y = 0.0011 * x + 0.4052$$

$$R = 1.0000$$

$$DL = 10.44$$

Weight: <None>

Min Conc: 0

24 Mg [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	413.91	0.0031	P	8.6
2	☐	500.000	503.821	38913.85	0.2897	P	3.3
3	☐	5000.000	5787.677	426310.44	3.2952	P	8.6
4	☐	12500.000	13450.009	1059164.45	7.6537	P	2.3
5	☐	25000.000	26493.343	2024284.44	15.0730	A	1.1
6	☐	50000.000	52961.176	3976237.76	30.1285	A	1.2
7	☐	100000.00	103258.17	8012953.04	58.7384	A	1.3
8	☐	500000.00	498945.94	39528621.72	283.8136	A	0.9


$$y = 5.6882E-004 * x + 0.0031$$

$$R = 1.0000$$

$$DL = 1.403$$

Weight: <None>

Min Conc: 0

27 Al [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	756.54	0.0057	P	10.4
2	☐	20.000	19.927	1291.81	0.0096	P	1.3
3	☐	1000.000	1079.669	28505.43	0.2203	P	7.9
4	☐	2500.000	2504.987	69689.30	0.5036	P	2.2
5	☐	5000.000	5053.963	135678.75	1.0103	P	0.9
6	☐	10000.000	10069.533	264916.45	2.0073	P	0.4
7	☐	20000.000	19821.501	538251.99	3.9457	P	0.2
8	☐	100000.00	100025.12	2769987.01	19.8884	A	0.8

$$y = 1.9878E-004 * x + 0.0057$$


$$R = 1.0000$$

$$DL = 8.89$$

Weight: <None>

Min Conc: 0

28 Si [No Gas] ISTD:45 Sc [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	851728.77	0.2795	P	3.3
2	☐	10.000	8.458	878275.86	0.2857	M	0.9
3	☐	50.000	43.197	977408.32	0.3111	A	0.9
4	☐	125.000	126.353	1174570.12	0.3720	A	1.6
5	☐	250.000	268.153	1521023.90	0.4759	A	0.7
6	☐	500.000	561.543	2164426.24	0.6908	A	3.7
7	☐	1000.000	964.877	3135079.58	0.9862	A	1.3
8	☐			1029643.38	0.2920	A	11.9

$$y = 7.3238\text{E-}004 * x + 0.2795$$


$$R = 0.9967$$

$$DL = 37.4$$

Weight: <None>

Min Conc: 0

31 P [He] ISTD:45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	4.611	227.78	0.0017	P	6.1
2	☐	0.000	-4.611	208.34	0.0016	P	10.1
3	☐	1000.000	1083.661	2422.50	0.0188	P	14.7
4	☐	2500.000	2535.911	5779.09	0.0418	P	5.0
5	☐	5000.000	5094.665	11048.92	0.0823	P	1.4
6	☐	10000.000	9974.469	21054.63	0.1595	P	1.9
7	☐	20000.000	19980.428	43368.29	0.3179	P	1.1
8	☐			263.90	0.0019	P	13.4

$$y = 1.5829\text{E-}005 * x + 0.0016$$


$$R = 1.0000$$

$$DL = 24.77$$

Weight: <None>

Min Conc: 0

34 S [No Gas] ISTD:45 Sc [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	-185.385	567866.88	0.1863	P	3.2
2	☐	0.000	185.385	578075.33	0.1880	P	1.8
3	☐	1000.000	1320.879	607157.25	0.1933	P	1.0
4	☐	2500.000	2596.251	628787.09	0.1992	P	0.9
5	☐	5000.000	4552.681	665266.39	0.2082	P	1.6
6	☐	10000.000	10784.368	742264.14	0.2369	P	1.2
7	☐	20000.000	19691.570	883349.94	0.2779	P	1.3
8	☐			549590.40	0.1559	P	12.6

$$y = 4.6065E-006 * x + 0.1872$$


$$R = 0.9983$$

$$DL = 3053$$

Weight: <None>

Min Conc: 0

39 K [He] ISTD:45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	40170.62	0.2997	P	0.7
2	☐	500.000	476.058	72649.69	0.5408	P	2.9
3	☐	2500.000	2842.658	225018.88	1.7391	P	8.4
4	☐	6250.000	6491.259	496357.81	3.5867	P	1.9
5	☐	12500.000	12968.621	922185.03	6.8665	P	0.9
6	☐	25000.000	25909.429	1771063.78	13.4193	A	0.5
7	☐	50000.000	50534.174	3531518.96	25.8883	A	1.4
8	☐	250000.00	249769.38	17655687.55	126.7735	A	0.7


$$y = 5.0636E-004 * x + 0.2997$$

$$R = 1.0000$$

$$DL = 12.8$$

Weight: <None>

Min Conc: 0

43 Ca [No Gas] ISTD:45 Sc [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	505.58	0.0002	P	8.6
2	☐	500.000	482.025	4809.28	0.0016	P	3.0
3	☐	5000.000	5536.591	50974.42	0.0162	P	1.0
4	☐	12500.000	13860.119	127452.28	0.0404	P	0.9
5	☐	25000.000	27514.615	255562.97	0.0800	P	2.0
6	☐	50000.000	54925.114	499687.07	0.1595	P	2.0
7	☐	100000.00	108593.24	1001742.66	0.3151	P	0.8
8	☐	500000.00	497623.75	5084766.35	1.4435	A	13.2


$$y = 2.9004\text{E-}006 * x + 1.6610\text{E-}004$$

$$R = 0.9998$$

$$DL = 14.75$$

Weight: <None>

Min Conc: 0

44 Ca [No Gas] ISTD:45 Sc [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	50778.01	0.0167	P	4.9
2	☐	500.000	513.295	126669.29	0.0412	P	2.1
3	☐	5000.000	5590.442	891879.17	0.2839	P	0.9
4	☐	12500.000	14129.394	2185193.10	0.6921	A	0.9
5	☐	25000.000	27504.821	4255619.03	1.3315	A	0.5
6	☐	50000.000	54863.628	8270980.25	2.6394	A	0.6
7	☐	100000.00	109563.46	16702272.97	5.2543	A	0.3
8	☐	500000.00	497429.05	83835591.34	23.7959	A	13.0

$$y = 4.7804\text{E-}005 * x + 0.0167$$


$$R = 0.9998$$

$$DL = 50.89$$

Weight: <None>

Min Conc: 0

47 Ti [No Gas] ISTD:45 Sc [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	536.14	0.0002	P	29.1
2	☐	5.000	4.563	2922.61	0.0010	P	4.7
3	☐	250.000	251.586	134784.42	0.0429	P	1.4
4	☐	625.000	624.579	335508.49	0.1063	P	1.1
5	☐	1250.000	1242.001	674778.71	0.2111	P	1.0
6	☐	2500.000	2511.111	1336953.14	0.4267	A	2.4
7	☐	5000.000	4996.418	2698531.83	0.8488	A	1.3
8	☐			1083.40	0.0003	P	21.2

$$y = 1.6985E-004 * x + 1.7583E-004$$


$$R = 1.0000$$

$$DL = 0.9039$$

Weight: <None>

Min Conc: 0

51 V [He] ISTD:45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	22.22	0.0002	P	22.3
2	☐	5.000	4.595	4684.15	0.0349	P	3.6
3	☐	50.000	53.517	52325.18	0.4043	P	7.7
4	☐	125.000	123.663	129254.27	0.9340	P	2.0
5	☐	250.000	248.264	251805.52	1.8749	P	0.1
6	☐	500.000	500.819	499160.15	3.7820	P	0.4
7	☐	1000.000	1000.018	1030160.11	7.5517	P	0.1
8	☐			324.45	0.0023	P	5.0


$$y = 0.0076 * x + 1.6562E-004$$

$$R = 1.0000$$

$$DL = 0.0147$$

Weight: <None>

Min Conc: 0

52 Cr [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	1355.63	0.0101	P	6.2
2	☐	2.000	1.939	3786.09	0.0282	P	1.5
3	☐	50.000	54.165	66541.61	0.5145	P	9.1
4	☐	125.000	123.671	160769.41	1.1616	P	1.7
5	☐	250.000	252.540	317162.48	2.3616	P	0.4
6	☐	500.000	505.151	622114.68	4.7137	P	0.4
7	☐	1000.000	996.748	1267408.61	9.2910	P	0.6
8	☐			6734.98	0.0484	P	1.2


$$y = 0.0093 * x + 0.0101$$

$$R = 1.0000$$

$$DL = 0.2019$$

Weight: <None>

Min Conc: 0

54 Fe [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	585.21	0.0044	P	12.2
2	☐	50.000	48.216	3104.10	0.0231	P	5.6
3	☐	2500.000	2827.041	142478.54	1.1014	P	8.7
4	☐	6250.000	6531.545	351331.89	2.5389	P	2.7
5	☐	12500.000	12981.275	677117.27	5.0417	P	0.4
6	☐	25000.000	26140.653	1339365.83	10.1482	P	0.8
7	☐	50000.000	52673.568	2788762.74	20.4441	A	1.1
8	☐	250000.00	249316.84	13474801.80	96.7508	A	0.2


$$y = 3.8805E-004 * x + 0.0044$$

$$R = 0.9999$$

$$DL = 4.118$$

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	123.34	0.0009	P	13.4
2	☐	1.000	0.944	686.69	0.0051	P	4.8
3	☐	500.000	543.560	311965.59	2.4109	P	8.2
4	☐	1250.000	1266.160	777043.25	5.6148	P	1.7
5	☐	2500.000	2542.530	1514180.93	11.2739	A	1.2
6	☐	5000.000	5070.295	2967149.64	22.4815	A	1.2
7	☐	10000.000	9950.022	6018032.06	44.1171	A	0.9
8	☐			5885.71	0.0423	P	3.3


$$y = 0.0044 * x + 9.2165E-004$$

$$R = 1.0000$$

$$DL = 0.08363$$

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	3709.81	0.0277	P	1.6
2	☐	50.000	49.563	51026.31	0.3798	P	2.5
3	☐	2500.000	2780.992	2560046.93	19.7835	A	8.1
4	☐	6250.000	6368.334	6264994.06	45.2676	A	2.1
5	☐	12500.000	12626.906	12050647.38	89.7278	A	0.4
6	☐	25000.000	25642.673	24046261.65	182.1902	A	0.7
7	☐	50000.000	49938.209	48395813.12	354.7829	A	1.0
8	☐	250000.00	249935.97	247279090.7	1,775.543	A	0.9


$$y = 0.0071 * x + 0.0277$$

$$R = 1.0000$$

$$DL = 0.1812$$

Weight: <None>

Min Conc: 0

57 Fe [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	188.89	0.0014	P	10.0
2	☐	50.000	47.741	1316.75	0.0098	P	5.4
3	☐	2500.000	2791.163	63677.08	0.4921	P	8.3
4	☐	6250.000	6442.129	156947.63	1.1340	P	1.2
5	☐	12500.000	12790.971	302212.82	2.2502	P	0.4
6	☐	25000.000	25574.673	593640.49	4.4978	P	0.6
7	☐	50000.000	50643.747	1214764.61	8.9053	P	1.0
8	☐	250000.00	249791.52	6116620.04	43.9182	A	0.4


$$y = 1.7581\text{E-}004 * x + 0.0014$$

$$R = 1.0000$$

$$DL = 2.408$$

Weight: <None>

Min Conc: 0

59 Co [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	46.67	0.0003	P	13.9
2	☐	1.000	0.968	2086.84	0.0155	P	2.5
3	☐	50.000	54.549	110782.39	0.8559	P	7.6
4	☐	125.000	127.649	277120.23	2.0023	P	1.2
5	☐	250.000	253.014	532979.22	3.9684	P	0.2
6	☐	500.000	508.982	1053596.07	7.9829	P	0.2
7	☐	1000.000	994.197	2127013.91	15.5927	A	0.7
8	☐			3999.50	0.0287	P	3.5


$$y = 0.0157 * x + 3.4805\text{E-}004$$

$$R = 0.9999$$

$$DL = 0.009287$$

Weight: <None>

Min Conc: 0

60 Ni [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	80.00	0.0006	P	25.3
2	☐	1.000	0.943	631.13	0.0047	P	2.7
3	☐	50.000	54.682	30826.74	0.2383	P	8.7
4	☐	125.000	126.759	76340.01	0.5516	P	1.8
5	☐	250.000	253.696	148183.47	1.1034	P	1.2
6	☐	500.000	505.041	289826.01	2.1960	P	0.4
7	☐	1000.000	996.101	590733.36	4.3306	P	0.9
8	☐			2472.46	0.0178	P	1.8


$$y = 0.0043 * x + 5.9679E-004$$

$$R = 1.0000$$

$$DL = 0.1043$$

Weight: <None>

Min Conc: 0

63 Cu [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	724.47	0.0054	P	13.1
2	☐	2.000	2.020	3843.90	0.0286	P	3.2
3	☐	500.000	554.648	825035.52	6.3760	P	8.2
4	☐	1250.000	1281.130	2037040.54	14.7203	A	2.8
5	☐	2500.000	2550.209	3934612.72	29.2967	A	1.1
6	☐	5000.000	5096.375	7726121.97	58.5415	A	1.3
7	☐	10000.000	9932.636	15562980.18	114.0900	A	1.0
8	☐			8720.53	0.0626	P	2.8


$$y = 0.0115 * x + 0.0054$$

$$R = 0.9999$$

$$DL = 0.1848$$

Weight: <None>

Min Conc: 0

66 Zn [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	208.89	0.0016	P	14.2
2	☐	2.000	1.915	648.91	0.0048	P	2.9
3	☐	500.000	543.396	120331.10	0.9294	P	6.9
4	☐	1250.000	1260.449	298051.16	2.1537	P	1.8
5	☐	2500.000	2515.673	577066.17	4.2969	P	1.1
6	☐	5000.000	5050.208	1138263.39	8.6244	P	0.2
7	☐	10000.000	9967.502	2321778.13	17.0203	A	1.2
8	☐			7101.84	0.0510	P	2.6


$$y = 0.0017 * x + 0.0016$$

$$R = 1.0000$$

$$DL = 0.3877$$

Weight: <None>

Min Conc: 0

75 As [He] ISTD:89 Y [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	12.33	0.0000	P	41.3
2	☐	1.000	0.935	182.35	0.0002	P	6.2
3	☐	50.000	53.928	9336.69	0.0086	P	8.2
4	☐	125.000	124.792	23180.26	0.0199	P	1.0
5	☐	250.000	251.381	45398.33	0.0401	P	0.3
6	☐	500.000	497.654	89606.37	0.0794	P	1.3
7	☐	1000.000	1000.657	185072.92	0.1596	P	0.5
8	☐			100.34	0.0001	P	2.5

$$y = 1.5945E-004 * x + 1.0883E-005$$


$$R = 1.0000$$

$$DL = 0.08466$$

Weight: <None>

Min Conc: 0

77 Se [He] ISTD:89 Y [He]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	5.55	0.0000	P	125.
2	☐	5.000	3.532	14.44	0.0000	P	11.1
3	☐	50.000	51.583	127.78	0.0001	P	16.5
4	☐	125.000	126.277	326.67	0.0003	P	18.1
5	☐	250.000	275.880	687.80	0.0006	P	6.6
6	☐	500.000	516.879	1280.07	0.0011	P	2.9
7	☐	1000.000	984.859	2500.25	0.0022	P	7.0
8	☐			3.33	0.0000	P	101.

$$y = 2.1840\text{E-}006 * x + 4.9280\text{E-}006$$


$$R = 0.9994$$

$$DL = 8.519$$

Weight: <None>

Min Conc: 0

78 Se [He] ISTD:89 Y [He]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	38.67	0.0000	P	13.7
2	☐	5.000	4.527	76.34	0.0001	P	13.7
3	☐	50.000	55.482	474.06	0.0004	P	10.6
4	☐	125.000	128.542	1128.16	0.0010	P	1.6
5	☐	250.000	257.370	2158.05	0.0019	P	2.2
6	☐	500.000	498.424	4132.03	0.0037	P	1.7
7	☐	1000.000	998.231	8460.06	0.0073	P	0.5
8	☐			60.01	0.0000	P	10.9


$$y = 7.2730\text{E-}006 * x + 3.4114\text{E-}005$$

$$R = 1.0000$$


$$DL = 1.932$$

Weight: <None>

Min Conc: 0

81 Br [No Gas] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			2235.83		P	8.0
2	☐			1972.17		P	6.8
3	☐			2078.97		P	4.1
4	☐			2062.28		P	11.4
5	☐			2088.98		P	15.9
6	☐			1958.82		P	7.8
7	☐			2189.10		P	3.2
8	☐			2699.71		P	6.5

82 Se [No Gas] ISTD: 89 Y [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	35.02	0.0000	P	126.
2	☐	5.000	4.623	326.10	0.0000	P	12.3
3	☐	50.000	49.770	3207.29	0.0004	P	2.3
4	☐	125.000	127.548	8091.94	0.0010	P	2.0
5	☐	250.000	252.803	16024.56	0.0019	P	1.0
6	☐	500.000	500.778	31746.93	0.0038	P	2.0
7	☐	1000.000	998.605	63339.52	0.0075	P	0.4
8	☐			29.13	0.0000	P	36.6


$$y = 7.4893E-006 * x + 4.3461E-006$$

$$R = 1.0000$$


$$DL = 2.203$$

Weight: <None>

Min Conc: 0

83 Kr [No Gas] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			318.90		P	18.4
2	☐			353.34		P	10.6
3	☐			335.56		P	0.6
4	☐			371.12		P	11.1
5	☐			372.23		P	11.9
6	☐			324.45		P	13.4
7	☐			403.34		P	6.8
8	☐			351.12		P	4.9

86 Sr [No Gas] ISTD : 89 Y [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	652.81	0.0001	P	14.1
2	☐	1.000	0.810	1577.91	0.0002	P	9.3
3	☐	50.000	50.020	57895.68	0.0068	P	0.5
4	☐	125.000	125.249	142718.45	0.0169	P	0.9
5	☐	250.000	250.020	284586.88	0.0337	P	1.3
6	☐	500.000	494.316	562732.39	0.0666	P	1.5
7	☐	1000.000	1002.805	1142162.05	0.1349	P	0.3
8	☐			7299.29	0.0008	P	16.0

$$y = 1.3448\text{E-}004 * x + 7.9294\text{E-}005$$


$$R = 1.0000$$

$$DL = 0.2493$$

Weight: <None>

Min Conc: 0

88 Sr [No Gas] ISTD: 89 Y [No Gas]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	58.34	0.0000	P	40.5
2	☐	1.000	0.942	9144.88	0.0011	P	4.7
3	☐	50.000	50.100	490787.41	0.0577	P	1.0
4	☐	125.000	126.605	1229596.28	0.1458	P	0.6
5	☐	250.000	256.132	2490609.58	0.2949	A	1.6
6	☐	500.000	503.044	4898210.79	0.5793	A	1.1
7	☐	1000.000	996.740	9715311.85	1.1478	A	0.4
8	☐			56980.74	0.0061	P	13.8

$$y = 0.0012 * x + 7.1461E-006$$


$$R = 1.0000$$

$$DL = 0.007531$$

Weight: <None>

Min Conc: 0

90 Zr [No Gas] ISTD: 89 Y [No Gas]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	333.35	0.0000	P	13.6
2	☐	1.000	0.934	5898.63	0.0007	P	3.4
3	☐	50.000	50.384	305079.05	0.0359	P	1.4
4	☐	125.000	126.556	759311.19	0.0900	P	0.5
5	☐	250.000	255.470	1534423.33	0.1817	A	0.5
6	☐	500.000	500.891	3011678.86	0.3562	A	2.6
7	☐	1000.000	997.973	6007327.96	0.7097	A	0.6
8	☐			2497.51	0.0003	P	10.2


$$y = 7.1109E-004 * x + 4.0525E-005$$

$$R = 1.0000$$

$$DL = 0.02318$$

Weight: <None>

Min Conc: 0

91 Zr [No Gas] ISTD:89 Y [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	75.00	0.0000	P	27.4
2	☐	1.000	0.933	1316.76	0.0002	P	0.8
3	☐	50.000	49.852	67399.28	0.0079	P	1.7
4	☐	125.000	127.205	170411.93	0.0202	P	0.8
5	☐	250.000	253.316	339707.61	0.0402	P	0.7
6	☐	500.000	497.197	667528.49	0.0789	P	1.3
7	☐	1000.000	1000.304	1344372.70	0.1588	P	0.6
8	☐			600.03	0.0001	P	5.7


$$y = 1.5877\text{E-}004 * x + 9.0891\text{E-}006$$

$$R = 1.0000$$

$$DL = 0.04709$$

Weight: <None>

Min Conc: 0

94 Mo [No Gas] ISTD:89 Y [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	183.34	0.0000	P	7.3
2	☐	5.000	4.605	6938.02	0.0008	P	2.9
3	☐	250.000	247.258	368259.50	0.0433	P	0.8
4	☐	625.000	627.509	926287.04	0.1098	P	1.6
5	☐	1250.000	1278.489	1889656.31	0.2238	A	1.0
6	☐	2500.000	2480.294	3670147.83	0.4341	A	1.7
7	☐	5000.000	5002.555	7409983.85	0.8755	A	1.5
8	☐			2822.58	0.0003	P	9.0


$$y = 1.7500\text{E-}004 * x + 2.2299\text{E-}005$$

$$R = 1.0000$$

$$DL = 0.02809$$

Weight: <None>

Min Conc: 0

95 Mo [No Gas] ISTD:89 Y [No Gas]

	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	36.11	0.0000	P	11.9
2	☐	5.000	4.691	8374.97	0.0010	P	0.8
3	☐	250.000	252.067	455116.27	0.0535	P	0.4
4	☐	625.000	630.846	1129120.52	0.1339	P	1.1
5	☐	1250.000	1277.477	2289555.50	0.2711	A	0.7
6	☐	2500.000	2506.279	4497356.79	0.5319	A	1.8
7	☐	5000.000	4989.157	8961862.61	1.0589	A	1.3
8	☐			3542.22	0.0004	P	11.4


$$y = 2.1223\text{E-}004 * x + 4.3913\text{E-}006$$

$$R = 1.0000$$

$$DL = 0.007362$$

Weight: <None>

Min Conc: 0

96 Mo [No Gas] ISTD:89 Y [No Gas]

	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	69.45	0.0000	P	22.9
2	☐	5.000	4.665	9292.23	0.0011	P	1.3
3	☐	250.000	250.445	502841.40	0.0591	P	1.0
4	☐	625.000	631.101	1256186.47	0.1490	P	1.2
5	☐	1250.000	1266.027	2523113.82	0.2988	A	1.8
6	☐	2500.000	2498.114	4984946.38	0.5896	A	1.4
7	☐	5000.000	4996.152	9980866.46	1.1791	A	0.2
8	☐			4145.18	0.0004	P	12.5

$$y = 2.3601\text{E-}004 * x + 8.4220\text{E-}006$$


$$R = 1.0000$$

$$DL = 0.02452$$

Weight: <None>

Min Conc: 0

97 Mo [No Gas] ISTD:89 Y [No Gas]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	19.45	0.0000	P	23.1
2	☐	5.000	4.488	5014.91	0.0006	P	2.3
3	☐	250.000	246.876	279075.71	0.0328	P	0.8
4	☐	625.000	625.631	701134.04	0.0831	P	1.2
5	☐	1250.000	1251.280	1404166.38	0.1663	M	0.9
6	☐	2500.000	2494.903	2803111.93	0.3315	A	1.8
7	☐	5000.000	5002.306	5626365.40	0.6647	A	0.7
8	☐			1994.64	0.0002	P	10.9

$$y = 1.3289\text{E-}004 * x + 2.3622\text{E-}006$$


$$R = 1.0000$$

$$DL = 0.01233$$

Weight: <None>

Min Conc: 0

98 Mo [No Gas] ISTD:89 Y [No Gas]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	47.22	0.0000	P	11.8
2	☐	5.000	4.511	12800.70	0.0015	P	3.5
3	☐	250.000	246.303	707497.56	0.0832	P	0.6
4	☐	625.000	636.534	1812613.06	0.2149	A	0.6
5	☐	1250.000	1248.288	3559215.75	0.4215	A	2.0
6	☐	2500.000	2484.210	7091922.21	0.8388	A	2.0
7	☐	5000.000	5007.066	14310274.28	1.6907	A	0.4
8	☐			5290.03	0.0006	P	6.4


$$y = 3.3766\text{E-}004 * x + 5.7565\text{E-}006$$

$$R = 1.0000$$

$$DL = 0.006027$$

Weight: <None>

Min Conc: 0

106 Cd [No Gas] ISTD :159 Tb [No Gas]

	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	2058.54	0.0005	P	9.9
2	☐	1.000	-0.209	2080.77	0.0005	P	7.1
3	☐	50.000	47.619	6337.72	0.0014	P	4.6
4	☐	125.000	124.876	13059.34	0.0029	P	1.0
5	☐	250.000	250.109	24080.02	0.0053	P	1.4
6	☐	500.000	506.204	46222.48	0.0103	P	1.9
7	☐	1000.000	997.007	90876.05	0.0198	P	1.6
8	☐			2325.27	0.0005	P	4.7


$$y = 1.9410\text{E-}005 * x + 4.7520\text{E-}004$$

$$R = 1.0000$$

$$DL = 7.293$$

Weight: <None>

Min Conc: 0

107 Ag [No Gas] ISTD :159 Tb [No Gas]

	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	66.67	0.0000	P	5.0
2	☐	1.000	0.953	4045.16	0.0009	P	3.4
3	☐	50.000	50.892	217604.67	0.0481	P	2.7
4	☐	125.000	127.052	540543.88	0.1200	P	1.4
5	☐	250.000	256.089	1092589.48	0.2419	P	2.2
6	☐	500.000	513.950	2178190.39	0.4854	A	1.2
7	☐	1000.000	991.201	4290639.55	0.9361	A	1.0
8	☐			2639.23	0.0005	P	15.0


$$y = 9.4439\text{E-}004 * x + 1.5378\text{E-}005$$

$$R = 0.9998$$

$$DL = 0.002463$$

Weight: <None>

Min Conc: 0

108 Cd [No Gas] ISTD :159 Tb [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	5.56	0.0000	P	111.
2	☐	1.000	1.204	79.17	0.0000	P	22.3
3	☐	50.000	50.398	3151.80	0.0007	P	6.4
4	☐	125.000	129.094	8030.19	0.0018	P	0.8
5	☐	250.000	252.749	15760.54	0.0035	P	2.7
6	☐	500.000	516.364	31981.55	0.0071	P	1.8
7	☐	1000.000	990.599	62653.74	0.0137	P	2.2
8	☐			63.89	0.0000	P	10.8


$$y = 1.3798\text{E-}005 * x + 1.2815\text{E-}006$$

$$R = 0.9998$$

$$DL = 0.3114$$

Weight: <None>

Min Conc: 0

109 Ag [No Gas] ISTD :159 Tb [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	36.11	0.0000	P	68.5
2	☐	1.000	0.936	3811.75	0.0009	P	9.3
3	☐	50.000	50.236	207651.42	0.0459	P	2.1
4	☐	125.000	125.856	517650.51	0.1149	P	1.0
5	☐	250.000	249.459	1028966.45	0.2278	P	1.9
6	☐	500.000	509.642	2088293.86	0.4653	A	1.2
7	☐	1000.000	995.195	4164742.37	0.9086	A	1.1
8	☐			2316.92	0.0005	P	10.7


$$y = 9.1302\text{E-}004 * x + 8.1339\text{E-}006$$

$$R = 0.9999$$

$$DL = 0.0183$$

Weight: <None>

Min Conc: 0

111 Cd [No Gas] ISTD :159 Tb [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	1448.20	0.0003	P	10.0
2	☐	1.000	0.944	2269.22	0.0005	P	6.5
3	☐	50.000	50.154	44673.07	0.0099	P	2.3
4	☐	125.000	128.508	111540.84	0.0248	P	2.1
5	☐	250.000	257.458	222631.93	0.0493	P	1.3
6	☐	500.000	505.856	433045.78	0.0965	P	0.5
7	☐	1000.000	994.761	868259.46	0.1894	P	1.8
8	☐			2094.10	0.0004	P	4.4


$$y = 1.9010E-004 * x + 3.3428E-004$$

$$R = 0.9999$$

$$DL = 0.5252$$

Weight: <None>

Min Conc: 0

118 Sn [No Gas] ISTD :159 Tb [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	213.90	0.0000	P	25.2
2	☐	5.000	4.706	11877.71	0.0027	P	0.1
3	☐	50.000	50.618	128681.23	0.0284	P	2.9
4	☐	125.000	126.345	319328.28	0.0709	P	1.7
5	☐	250.000	252.057	638728.68	0.1414	P	1.6
6	☐	500.000	505.028	1270997.35	0.2832	P	1.1
7	☐	1000.000	996.774	2562044.22	0.5590	A	0.9
8	☐			2819.81	0.0006	P	9.9

$$y = 5.6071E-004 * x + 4.9532E-005$$


$$R = 1.0000$$

$$DL = 0.0667$$

Weight: <None>

Min Conc: 0

121 Sb [No Gas] ISTD :159 Tb [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	27.78	0.0000	P	14.2
2	☐	2.000	1.886	7071.48	0.0016	P	1.7
3	☐	50.000	49.950	191114.28	0.0422	P	2.8
4	☐	125.000	124.761	475003.00	0.1055	P	1.9
5	☐	250.000	251.097	958894.31	0.2122	P	1.5
6	☐	500.000	511.635	1940733.94	0.4325	A	0.7
7	☐	1000.000	993.941	3850815.06	0.8401	A	0.7
8	☐			2964.30	0.0006	P	11.7

$$y = 8.4524\text{E-}004 * x + 6.3827\text{E-}006$$


$$R = 0.9999$$

$$DL = 0.003213$$

Weight: <None>

Min Conc: 0

135 Ba [No Gas] ISTD :159 Tb [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	11.11	0.0000	P	111.
2	☐	10.000	9.634	6824.15	0.0015	P	1.4
3	☐	250.000	248.449	180069.70	0.0398	P	1.0
4	☐	625.000	633.207	456588.45	0.1014	P	1.3
5	☐	1250.000	1268.234	917092.78	0.2030	P	2.5
6	☐	2500.000	2520.842	1810952.73	0.4035	A	0.4
7	☐	5000.000	4984.073	3656760.63	0.7978	A	1.9
8	☐			7185.44	0.0014	P	13.6


$$y = 1.6008\text{E-}004 * x + 2.5264\text{E-}006$$

$$R = 1.0000$$

$$DL = 0.0526$$

Weight: <None>

Min Conc: 0

137 Ba [No Gas] ISTD : 159 Tb [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	22.22	0.0000	P	111.
2	☐	10.000	9.337	11413.46	0.0026	P	3.7
3	☐	250.000	250.833	313530.50	0.0693	P	1.5
4	☐	625.000	635.532	790399.86	0.1755	P	1.3
5	☐	1250.000	1285.927	1603795.48	0.3550	A	2.9
6	☐	2500.000	2516.883	3118339.00	0.6949	A	1.7
7	☐	5000.000	4981.220	6303616.98	1.3753	A	1.3
8	☐			12002.95	0.0024	P	12.2


$$y = 2.7609E-004 * x + 5.0537E-006$$

$$R = 1.0000$$


$$DL = 0.061$$

Weight: <None>


Min Conc: 0

150 Nd [No Gas] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			0.00		P	
2	☐			2.22		P	173.
3	☐			26.67		P	43.3
4	☐			17.78		P	21.7
5	☐			44.44		P	17.3
6	☐			97.78		P	10.4
7	☐			104.45		P	22.4
8	☐			22.22		P	45.8

150 Nd [He] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			0.00		P	
2	☐			0.00		P	
3	☐			3.33		P	100.
4	☐			13.33		P	66.2
5	☐			3.33		P	100.
6	☐			13.33		P	NaN
7	☐			23.33		P	14.3
8	☐			7.78		P	24.7

156 Dy [No Gas] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			13.89		P	91.6
2	☐			11.11		P	114.
3	☐			22.22		P	114.
4	☐			22.22		P	21.6
5	☐			44.44		P	43.3
6	☐			36.11		P	58.1
7	☐			94.45		P	43.5
8	☐			236.12		P	5.4

156 Dy [He] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			1.11		P	173.
2	☐			4.44		P	114.
3	☐			4.44		P	43.4
4	☐			5.56		P	34.7
5	☐			18.89		P	10.2
6	☐			20.00		P	44.1
7	☐			38.89		P	17.8
8	☐			111.11		P	25.0

160 Gd [No Gas] Noavailable calibration poin...


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			44.45		P	10.8
2	☐			19.44		P	99.0
3	☐			25.00		P	33.3
4	☐			52.78		P	55.5
5	☐			55.56		P	22.9
6	☐			55.56		P	67.6
7	☐			125.00		P	34.6
8	☐			244.45		P	32.2

160 Gd [He] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			18.89		P	36.7
2	☐			11.11		P	17.3
3	☐			24.45		P	64.4
4	☐			18.89		P	20.4
5	☐			22.22		P	45.8
6	☐			43.33		P	46.8
7	☐			48.89		P	37.6
8	☐			155.56		P	26.8

164 Er [No Gas] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			19.44		P	65.5
2	☐			13.89		P	69.3
3	☐			38.89		P	44.6
4	☐			36.11		P	26.7
5	☐			22.22		P	43.3
6	☐			30.55		P	15.7
7	☐			27.78		P	75.5
8	☐			61.11		P	15.7

164 Er [He] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			5.56		P	34.7
2	☐			11.11		P	17.3
3	☐			11.11		P	45.8
4	☐			6.67		P	50.0
5	☐			7.78		P	24.7
6	☐			12.22		P	31.5
7	☐			10.00		P	88.2
8	☐			17.78		P	10.8

203 Tl [No Gas] ISTD :209 Bi [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	19.45	0.0000	P	21.6
2	☐	1.000	0.914	1889.07	0.0008	P	2.6
3	☐	50.000	47.805	101307.62	0.0425	P	2.9
4	☐	125.000	120.745	254872.62	0.1074	P	0.3
5	☐	250.000	243.109	519166.54	0.2162	P	1.3
6	☐	500.000	482.664	1022806.97	0.4291	P	1.4
7	☐	1000.000	1011.032	2128307.69	0.8989	A	0.3
8	☐			1313.99	0.0005	P	3.4

$$y = 8.8909\text{E-}004 * x + 8.6404\text{E-}006$$


$$R = 0.9998$$

$$DL = 0.006293$$

Weight: <None>

Min Conc: 0

205 Tl [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	41.67	0.0000	P	57.6
2	☐	1.000	0.859	4153.56	0.0018	P	2.1
3	☐	50.000	48.682	241498.62	0.1013	P	2.2
4	☐	125.000	122.223	603853.44	0.2544	P	0.6
5	☐	250.000	243.339	1216366.47	0.5064	P	0.8
6	☐	500.000	494.823	2454413.27	1.0297	A	0.3
7	☐	1000.000	1004.667	4950288.35	2.0907	A	1.3
8	☐			3536.69	0.0014	P	12.0

$$y = 0.0021 * x + 1.8900E-005$$


$$R = 1.0000$$

$$DL = 0.0157$$

Weight: <None>

Min Conc: 0

206 [Pb] [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	481.50	0.0002	P	11.2
2	☐	1.000	0.821	1896.47	0.0008	P	9.1
3	☐	250.000	239.784	424119.89	0.1780	P	2.3
4	☐	625.000	606.412	1067553.86	0.4497	P	0.9
5	☐	1250.000	1249.278	2224638.91	0.9262	A	1.6
6	☐	2500.000	2497.583	4413030.62	1.8515	A	1.1
7	☐	5000.000	5004.223	8783020.34	3.7095	A	1.2
8	☐			4484.20	0.0018	P	9.9

$$y = 7.4123E-004 * x + 2.1543E-004$$


$$R = 1.0000$$

$$DL = 0.09723$$

Weight: <None>

Min Conc: 0

207 [Pb] [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	409.28	0.0002	P	4.9
2	☐	1.000	0.867	1761.26	0.0008	P	3.8
3	☐	250.000	237.806	381567.62	0.1601	P	2.8
4	☐	625.000	599.999	958349.73	0.4037	P	0.5
5	☐	1250.000	1251.129	2021406.09	0.8416	A	1.3
6	☐	2500.000	2487.065	3986922.95	1.6728	A	1.8
7	☐	5000.000	5009.920	7978230.04	3.3695	A	0.9
8	☐			3902.52	0.0016	P	9.6

$$y = 6.7253E-004 * x + 1.8275E-004$$


$$R = 1.0000$$

$$DL = 0.04003$$

Weight: <None>

Min Conc: 0

208 Pb [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	2001.96	0.0009	P	11.0
2	☐	1.000	0.805	7613.97	0.0033	P	3.2
3	☐	250.000	240.329	1720323.04	0.7218	P	2.3
4	☐	625.000	619.285	4412085.36	1.8586	A	0.4
5	☐	1250.000	1252.023	9022915.13	3.7566	A	1.0
6	☐	2500.000	2483.727	17760036.08	7.4513	A	0.8
7	☐	5000.000	5008.828	35576658.67	15.0258	A	0.9
8	☐			17994.13	0.0072	P	8.8

$$y = 0.0030 * x + 8.9524E-004$$


$$R = 1.0000$$

$$DL = 0.09862$$

Weight: <None>

Min Conc: 0

238 U [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	5.55	0.0000	P	86.7
2	☐	1.000	0.902	5017.79	0.0022	P	6.7
3	☐	50.000	48.428	278823.34	0.1170	P	1.5
4	☐	125.000	123.351	707239.06	0.2979	P	0.9
5	☐	250.000	244.279	1417162.79	0.5900	P	0.8
6	☐	500.000	500.513	2881325.37	1.2089	A	0.8
7	☐	1000.000	1001.459	5726965.06	2.4188	A	0.5
8	☐			1291.77	0.0005	P	16.3

$$y = 0.0024 * x + 2.5229E-006$$


$$R = 1.0000$$

$$DL = 0.002718$$


Weight: <None>

Min Conc: 0


6 Li (ISTD) [No Gas]




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	1.000		1215770.18		A	4.0
2	☐	1.000		1205406.97		A	2.2
3	☐	1.000		1253626.14		A	2.1
4	☐	1.000		1279834.06		A	2.1
5	☐	1.000		1304957.00		A	0.5
6	☐	1.000		1301024.19		A	0.8
7	☐	1.000		1263018.89		A	1.8
8	☐	1.000		1397123.61		A	7.2

45 Sc (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		3049863.42		A	3.7
2	☐	1.000		3074439.85		A	1.1
3	☐	1.000		3141612.17		A	1.4
4	☐	1.000		3157416.62		A	0.8
5	☐	1.000		3196085.96		A	1.2
6	☐	1.000		3133805.54		A	1.1
7	☐	1.000		3178849.78		A	0.9
8	☐	1.000		3561327.31		A	12.4

45 Sc (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		134034.03		P	1.8
2	☐	1.000		134403.36		P	2.4
3	☐	1.000		129902.61		P	7.2
4	☐	1.000		138415.23		P	1.5
5	☐	1.000		134303.91		P	0.6
6	☐	1.000		131981.40		P	0.6
7	☐	1.000		136414.60		P	0.6
8	☐	1.000		139272.74		P	0.5

89 Y (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		8215260.16		A	2.5
2	☐	1.000		8374773.48		A	1.2
3	☐	1.000		8506727.23		A	1.3
4	☐	1.000		8433634.25		A	1.0
5	☐	1.000		8445134.25		A	1.5
6	☐	1.000		8455997.72		A	1.2
7	☐	1.000		8464381.33		A	1.1
8	☐	1.000		9519990.34		A	12.0

89 Y (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		1134155.64		P	1.4
2	☐	1.000		1140677.45		P	2.5
3	☐	1.000		1088622.99		P	7.0
4	☐	1.000		1164317.54		P	0.6
5	☐	1.000		1132291.20		P	0.2
6	☐	1.000		1129148.66		P	0.9
7	☐	1.000		1159830.78		P	0.2
8	☐	1.000		1203883.56		P	1.6

103 Rh (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		4210377.35		A	3.4
2	☐	1.000		4238217.80		A	1.5
3	☐	1.000		4363864.38		A	0.8
4	☐	1.000		4286584.34		A	0.5
5	☐	1.000		4297169.19		A	1.1
6	☐	1.000		4272412.68		A	0.8
7	☐	1.000		4326794.50		A	1.7
8	☐	1.000		4628594.85		A	10.8

103 Rh (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		2322945.38		A	1.6
2	☐	1.000		2333098.99		A	2.5
3	☐	1.000		2212451.79		A	8.2
4	☐	1.000		2354883.73		A	1.9
5	☐	1.000		2294585.07		A	1.9
6	☐	1.000		2266777.28		A	1.2
7	☐	1.000		2297100.42		A	0.8
8	☐	1.000		2294986.02		A	1.4

115 In (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		4181946.19		A	3.9
2	☐	1.000		4295622.60		A	1.2
3	☐	1.000		4342401.91		A	1.2
4	☐	1.000		4363608.23		A	0.6
5	☐	1.000		4337399.38		A	2.3
6	☐	1.000		4277156.62		A	0.3
7	☐	1.000		4340836.70		A	2.1
8	☐	1.000		4749855.91		A	11.5

115 In (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		827292.96		P	1.0
2	☐	1.000		827303.61		P	2.1
3	☐	1.000		796660.03		P	6.9
4	☐	1.000		847810.20		P	1.4
5	☐	1.000		824191.88		P	0.2
6	☐	1.000		819031.15		P	0.4
7	☐	1.000		830807.16		P	0.2
8	☐	1.000		856942.44		P	1.0

159 Tb (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		4342548.89		A	4.9
2	☐	1.000		4417943.30		A	0.7
3	☐	1.000		4528079.17		A	2.4
4	☐	1.000		4504962.36		A	1.4
5	☐	1.000		4518904.69		A	2.3
6	☐	1.000		4487787.26		A	0.9
7	☐	1.000		4583734.00		A	0.8
8	☐	1.000		5084302.84		A	11.4

159 Tb (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		2376122.53		A	1.7
2	☐	1.000		2381671.91		A	2.4
3	☐	1.000		2260657.61		A	6.6
4	☐	1.000		2421674.55		A	1.2
5	☐	1.000		2383104.62		A	1.5
6	☐	1.000		2416786.30		A	1.2
7	☐	1.000		2480486.21		A	0.8
8	☐	1.000		2577088.68		A	1.2

165 Ho (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		4054908.68		A	4.6
2	☐	1.000		4102637.55		A	0.9
3	☐	1.000		4187729.65		A	2.1
4	☐	1.000		4227835.77		A	1.3
5	☐	1.000		4245278.00		A	0.5
6	☐	1.000		4250595.88		A	1.0
7	☐	1.000		4295449.71		A	1.8
8	☐	1.000		4741647.59		A	12.2

165 Ho (ISTD) [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		2336655.31		A	0.1
2	☐	1.000		2367795.17		A	1.5
3	☐	1.000		2288416.20		A	6.7
4	☐	1.000		2456984.56		A	1.5
5	☐	1.000		2392374.72		A	0.9
6	☐	1.000		2401216.40		A	0.8
7	☐	1.000		2475073.38		A	1.5
8	☐	1.000		2537304.74		A	1.8

209 Bi (ISTD) [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		2242423.67		A	4.4
2	☐	1.000		2300169.20		A	1.0
3	☐	1.000		2384414.89		A	2.9
4	☐	1.000		2374009.76		A	1.3
5	☐	1.000		2402140.30		A	1.4
6	☐	1.000		2383538.94		A	0.7
7	☐	1.000		2367681.14		A	0.8
8	☐	1.000		2529552.10		A	11.4

209 Bi (ISTD) [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		1541633.81		P	0.7
2	☐	1.000		1555644.18		P	1.2
3	☐	1.000		1506234.68		P	6.2
4	☐	1.000		1601835.65		M	1.4
5	☐	1.000		1568754.19		P	0.3
6	☐	1.000		1575604.86		P	0.7
7	☐	1.000		1595354.23		P	0.7
8	☐	1.000		1596070.90		P	1.2

US EPA Tune Check Report

Operator Name Jaswal
Acq/Data Batch D:\Agilent\ICPMH\1\DATA\P8081320-MS.b
Acq. Date-Time 2020-08-13 15:01:17
Report Comment ---
Instrument Name G8403A SG19224459

[No Gas]

Sensitivity

Mass	Conc. [ug/l]	Count	CPS	Resp (Required) [cps/ug/l]	Resp (Flag)	RSD%	RSD% (Required)
9		2710	27098.85			0.923	5.000
24		20864	208644.71			0.562	5.000
25		2752	27523.80			0.837	5.000
26		3235	32352.41			0.454	5.000
59		17333	173325.68			0.507	5.000
113		2521	25207.98			0.710	5.000
115		8371	83714.11			2.071	5.000
206		2384	23836.26			1.270	5.000
207		1955	19551.88			1.014	5.000
208		4907	49067.08			1.206	5.000
220		1	9.10			17.973	

Mass	RSD% (Flag)
9	
24	
25	
26	
59	
113	
115	
206	
207	
208	
220	


Mass	Rep#1 Count	Rep#2 Count	Rep#3 Count	Rep#4 Count	Rep#5 Count
9	2707	2752	2690	2707	2694
24	20795	21065	20771	20865	20827
25	2734	2788	2749	2760	2730
26	3216	3250	3238	3248	3224
59	17184	17393	17321	17390	17375
113	2511	2549	2514	2527	2503
115	8532	8550	8360	8275	8140
206	2350	2406	2384	2421	2358
207	1927	1982	1951	1956	1961
208	4819	4960	4937	4944	4874

US EPA Tune Check Report

Mass	Rep#1 Count	Rep#2 Count	Rep#3 Count	Rep#4 Count	Rep#5 Count
220	1	1	1	1	1

Integration Time [sec] 0.1

Resolution/Axis



Mass	Peak Height	Axis	Axis (Required)	Axis (Flag)
9	4767.55	9.00	8.90 - 9.10	
24	35786.74	23.95	23.90 - 24.10	
25	4728.41	25.00	24.90 - 25.10	
26	5584.87	26.00	25.90 - 26.10	
59	30869.57	58.95	58.90 - 59.10	
113	4949.93	113.00	112.90 - 113.10	
115	16100.60	115.00	114.90 - 115.10	
206	4540.93	206.00	205.90 - 206.10	
207	3737.74	207.00	206.90 - 207.10	
208	9236.71	208.00	207.90 - 208.10	
220			-	

Mass	W-50%	W-5%	W-5% (Required)	W-5% (Flag)
9	0.60	0.739	0.900	
24	0.62	0.813	0.900	
25	0.62	0.776	0.900	
26	0.62	0.825	0.900	
59	0.59	0.779	0.900	
113	0.53	0.731	0.900	
115	0.54	0.765	0.900	
206	0.55	0.788	0.900	
207	0.54	0.750	0.900	
208	0.55	0.816	0.900	
220				

Integration Time [sec] 0.1

Acquisition Time [sec] 256.770000000002

Y Axis Linear

Tune Parameters

Plasma Parameters

Plasma Mode HMI Nebulizer Gas 0.50 L/min Dilution Gas 0.45 L/min

US EPA Tune Check Report

RF Power	1600 W	Option Gas	---	Auxiliary Gas	0.90 L/min
RF Matching	1.40 V	Nebulizer Pump	0.10 rps	Plasma Gas	15.0 L/min
Sample Depth	10.0 mm	S/C Temp	2 °C		

Lens Parameters

Extract 1	-6.0 V	Omega Lens	10.4 V	Deflect	14.4 V
Extract 2	-245.0 V	Cell Entrance	-30 V	Plate Bias	-50 V
Omega Bias	-120 V	Cell Exit	-50 V		

Cell Parameters

Use Gas	No	3rd Gas Flow	---	Energy Discrimination	4.0 V
He Flow	0.0 mL/min	OctP Bias	-8.0 V		
H2 Flow	---	OctP RF	170 V		

QP Parameters

Mass Gain	123	Axis Gain	0.9994	QP Bias	-4.0 V
Mass Offset	124	Axis Offset	-0.03		

Hardware Settings

Torch

Torch H	0.2 mm	Torch V	-0.2 mm
---------	--------	---------	---------

EM

Discriminator	3.9 mV	Analog HV	2184 V	Pulse HV	1276 V
---------------	--------	-----------	--------	----------	--------

[He]

Sensitivity

Mass	Conc. [ug/l]	Count	CPS	Resp (Required) [cps/ug/l]	Resp (Flag)	RSD%	RSD% (Required)
59		4596	45957.15			1.611	
89		9139	91394.90			1.273	
205		7352	73519.05			1.276	

Mass	RSD% (Flag)
59	
89	
205	

Mass	Rep#1 Count	Rep#2 Count	Rep#3 Count	Rep#4 Count	Rep#5 Count
59	4470	4620	4665	4599	4624
89	8963	9254	9225	9095	9160
205	7218	7457	7405	7300	7380

Integration Time [sec] 0.1

Tune Parameters

Plasma Parameters

Plasma Mode	HMI	Nebulizer Gas	0.50 L/min	Dilution Gas	0.45 L/min
RF Power	1600 W	Option Gas	---	Auxiliary Gas	0.90 L/min
RF Matching	1.40 V	Nebulizer Pump	0.10 rps	Plasma Gas	15.0 L/min
Sample Depth	10.0 mm	S/C Temp	2 °C		

Lens Parameters

Extract 1	-6.0 V	Omega Lens	10.4 V	Deflect	4.4 V
-----------	--------	------------	--------	---------	-------

US EPA Tune Check Report

Extract 2	-245.0 V	Cell Entrance	-50 V	Plate Bias	-60 V
Omega Bias	-120 V	Cell Exit	-70 V		
Cell Parameters					
Use Gas	Yes	3rd Gas Flow	---	Energy Discrimination	5.0 V
He Flow	4.5 mL/min	OctP Bias	-18.0 V		
H2 Flow	---	OctP RF	200 V		
QP Parameters					
Mass Gain	123	Axis Gain	0.9994	QP Bias	-13.0 V
Mass Offset	124	Axis Offset	-0.03		
Hardware Settings					
Torch					
Torch H	0.2 mm	Torch V	-0.2 mm		
EM					
Discriminator	3.9 mV	Analog HV	2184 V	Pulse HV	1276 V