



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110526

Slope : 97.6

pH Meter ID : WC OH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.00	08/14/2020	09:25
2	CAL2	1	Water	NA	NA	21.1	7.01	08/14/2020	09:26
3	CAL3	1	Water	NA	NA	21.1	10.03	08/14/2020	09:27
4	ICV	1	Water	NA	NA	21.1	6.99	08/14/2020	09:28
5	CCV1	1	Water	NA	NA	21.1	2.02	08/14/2020	09:30
6	L3680-01	1	Water	NA	NA	25.0	10.70	08/14/2020	09:31
7	L3680-01DUP	1	Water	NA	NA	25.1	10.71	08/14/2020	09:32
8	CCV2	1	Water	NA	NA	21.1	12.02	08/14/2020	09:35

28 110526

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3680 WorkList ID : 141470 Department : Wet-Chemistry Date : 08-14-2020 09:18:28

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/13/2020	Water	L3680-01	pH	Cool 4 deg C	PSEG01	K21	DRUM-1-8	08/13/2020	9040C

Date/Time 08.14.2020 09:20
Received by: JB / wet-chem
Relinquished by: JB / wet-chem

Date/Time 08.14.2020 10:00
Received by: JB / wet-chem
Relinquished by: JB / wet-chem