



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB110527

Slope : 97.6

pH Meter ID : WC OH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.00	08/14/2020	09:26
2	CAL2	1	Water	NA	NA	21.1	7.01	08/14/2020	09:26
3	CAL3	1	Water	NA	NA	21.1	10.03	08/14/2020	09:27
4	ICV	1	Water	NA	NA	21.1	6.99	08/14/2020	09:28
5	CCV1	1	Water	NA	NA	21.2	2.02	08/14/2020	09:30
6	L3680-02	1	Water	NA	NA	25.0	10.70	08/14/2020	09:31
7	L3680-02DUP	1	Water	NA	NA	25.1	10.71	08/14/2020	09:32
8	CCV2	1	Water	NA	NA	21.1	12.02	08/14/2020	09:35

LB 110527

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity 13680

WorkList ID : 141471

Department : Wet-Chemistry

Date : 08-14-2020 09:19:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/13/2020	Water	L3680-02	Corrosivity	Cool 4 deg C	PSEG01	K21	DRUM-1-8	08/13/2020	9040C

Date/Time 08-14-2020 9:20
 Received by: Jo | wet
 Relinquished by: Jo | Galvin

Date/Time 08-14-2020 10:00
 Received by: Jo | wet
 Relinquished by: Jo | Galvin