


LB110528

7470A

INSTRUMENT ID: CV1

Linear ▼



A= 0.0000e+000

B= 3.2964e-004

C= 6.5897e-002

Rho= 0.9998312

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.058	0.058	-23	0.000	-23				
0.2	0.200	0.223	0.023	477	0.0 %	477				
2.5	2.500	2.361	-0.139	6964	0.0 %	6964				
5.0	5.000	5.022	0.022	15035	0.0 %	15035				
7.5	7.500	7.559	0.059	22730	0.0 %	22730				
10.0	10.000	9.976	-0.024	30065	0.0 %	30065				

LB110528 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 14 Aug 2020 10:12:55

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-23	-	0.00007470A	PPB		14 Aug 2020 10:34:38	S	Std
0.2 - 1	50.2	477	-	0.20007470A	PPB		14 Aug 2020 10:36:46	S	Std
2.5 - 1	50.5	6964	-	2.50007470A	PPB		14 Aug 2020 10:38:54	S	Std
5.0 - 1	55	15035	-	5.00007470A	PPB		14 Aug 2020 10:41:02	S	Std
7.5 - 1	57.5	22730	-	7.50007470A	PPB		14 Aug 2020 10:43:10	S	Std
10.0 - 1	510	30065	-	10.00007470A	PPB		14 Aug 2020 10:47:52	S	Std
ICV68 - 1	ICV68	12054	4.0394	-7470A	PPB		14 Aug 2020 10:50:21	U	SMPL
ICB68 - 1	ICB68	-287	-0.0287	-7470A	PPB		14 Aug 2020 10:52:28	U	SMPL
CCV18 - 1	CCV18	16082	5.3671	-7470A	PPB		14 Aug 2020 10:54:35	U	SMPL
CCB18 - 1	CCB18	-280	-0.0264	-7470A	PPB		14 Aug 2020 10:56:40	U	SMPL
CRA - 1	CRA	434	0.2090	-7470A	PPB		14 Aug 2020 10:58:48	U	SMPL
HighStd - 1	HighStd	30630	10.1627	-7470A	PPB		14 Aug 2020 11:00:53	U	SMPL
ChkStd - 1	ChkStd	22487	7.4785	-7470A	PPB		14 Aug 2020 11:02:59	U	SMPL
PB130952BL - 1	PBW	-381	-0.0597	-7470A	PPB		14 Aug 2020 11:05:08	U	SMPL
PB130952BS - 1	LCSW	11895	3.9869	-7470A	PPB		14 Aug 2020 11:07:16	U	SMPL
L3501-05 - 1	NB-08-081220	-40	0.0527	-7470A	PPB		14 Aug 2020 11:14:06	U	SMPL
L3501-05DUP - 1	NB-08-081220DUP	-60	0.0461	-7470A	PPB		14 Aug 2020 11:16:13	U	SMPL
L3501-05MS - 1	NB-08-081220MS	11802	3.9563	-7470A	PPB		14 Aug 2020 11:18:21	U	SMPL
L3501-05MSD - 1	NB-08-081220MSD	12742	4.2661	-7470A	PPB		14 Aug 2020 11:23:05	U	SMPL
L3508-16 - 1	TR-04-081220	-254	-0.0178	-7470A	PPB		14 Aug 2020 11:25:11	U	SMPL
CCV19 - 1	CCV19	16066	5.3619	-7470A	PPB		14 Aug 2020 11:27:19	U	SMPL
CCB19 - 1	CCB19	-262	-0.0205	-7470A	PPB		14 Aug 2020 11:29:24	U	SMPL
L3659-02 - 1	SP1-1	-78	0.0402	-7470A	PPB		14 Aug 2020 11:31:31	U	SMPL
L3659-05 - 1	SP1-2	-56	0.0474	-7470A	PPB		14 Aug 2020 11:33:37	U	SMPL
L3659-08 - 1	SP1-3	-49	0.0497	-7470A	PPB		14 Aug 2020 11:35:43	U	SMPL
L3660-02 - 1	RBR-57	-26	0.0573	-7470A	PPB		14 Aug 2020 11:37:51	U	SMPL
L3660-04 - 1	7211979	-26	0.0573	-7470A	PPB		14 Aug 2020 11:39:58	U	SMPL
L3660-06 - 1	VB25984	-27	0.0570	-7470A	PPB		14 Aug 2020 11:42:05	U	SMPL
L3660-08 - 1	RBR-58	9	0.0689	-7470A	PPB		14 Aug 2020 11:44:13	U	SMPL
L3660-10 - 1	RB47183	-22	0.0586	-7470A	PPB		14 Aug 2020 11:46:21	U	SMPL
L3662-02 - 1	VB16182	-11	0.0623	-7470A	PPB		14 Aug 2020 11:48:28	U	SMPL
L3662-04 - 1	RBR-374	1	0.0662	-7470A	PPB		14 Aug 2020 11:50:34	U	SMPL
CCV20 - 1	CCV20	15695	5.2396	-7470A	PPB		14 Aug 2020 11:52:41	U	SMPL
CCB20 - 1	CCB20	-197	0.0010	-7470A	PPB		14 Aug 2020 11:54:47	U	SMPL
L3662-06 - 1	RBR-59	-39	0.0530	-7470A	PPB		14 Aug 2020 11:56:54	U	SMPL
L3662-08 - 1	24051	-6	0.0639	-7470A	PPB		14 Aug 2020 11:59:00	U	SMPL
L3662-10 - 1	N48765	-9	0.0629	-7470A	PPB		14 Aug 2020 12:01:07	U	SMPL
L3677-03 - 1	TP-6	26	0.0745	-7470A	PPB		14 Aug 2020 12:03:13	U	SMPL
L3677-06 - 1	TP-1	-8	0.0633	-7470A	PPB		14 Aug 2020 12:05:20	U	SMPL
PB130952TB - 1	PB130935TB	-15	0.0610	-7470A	PPB		14 Aug 2020 12:07:27	U	SMPL
L3501-05LX5 - 1		-4	0.0646	-7470A	PPB		14 Aug 2020 12:09:34	U	SMPL
L3501-05A - 1		12057	4.0403	-7470A	PPB		14 Aug 2020 12:11:42	U	SMPL
CCV21 - 1	CCV21	15666	5.2300	-7470A	PPB		14 Aug 2020 12:13:49	U	SMPL
CCB21 - 1	CCB21	-231	-0.0102	-7470A	PPB		14 Aug 2020 12:15:56	U	SMPL