



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB110536

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	08/14/2020	09:45
2	L3680-02		1	8	100.00	>212.0	>212.0	08/14/2020	10:10
3	L3680-02DUP		1	8	100.00	>212.0	>212.0	08/14/2020	10:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB110536

WorkList Name : fp81420

WorkList ID : 141468

Department : Wet-Chemistry

Date : 08-14-2020 09:10:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/19/2020	Water	L3680-02	Flash Point	Cool 4 deg C	PSEG01	K21	DRUM-1-8	08/13/2020	1010A

Date/Time 08/14/2020 09:15
 Received by: KS (wekchen)
 Relinquished by: KS (wekchen)

Date/Time 08/14/2020 10:45
 Received by: KS (wekchen)
 Relinquished by: KS (wekchen)