



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110539

Slope : 97.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	3.99	08/14/2020	14:00
2	CAL2	1	Water	NA	NA	21.1	7.00	08/14/2020	14:01
3	CAL3	1	Water	NA	NA	21.1	10.03	08/14/2020	14:02
4	ICV	1	Water	NA	NA	21.1	7.01	08/14/2020	14:05
5	CCV1	1	Water	NA	NA	21.1	2.02	08/14/2020	14:06
6	L3690-01	1	Water	NA	NA	26.0	6.24	08/14/2020	14:10
7	L3690-02	1	Water	NA	NA	26.2	6.22	08/14/2020	14:11
8	L3690-03	1	Water	NA	NA	25.8	6.47	08/14/2020	14:13
9	L3690-03DUP	1	Water	NA	NA	25.9	6.48	08/14/2020	14:14
10	CCV2	1	Water	NA	NA	21.2	12.02	08/14/2020	14:15

LB 116539

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3690 WorkList ID : 141482 Department : Wet-Chemistry Date : 08-14-2020 13:44:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/14/2020	Water	L3690-01	pH	Cool 4 deg C	PSEG01	K22	NWB-1518	08/14/2020	9040C
08/14/2020	Water	L3690-02	pH	Cool 4 deg C	PSEG01	K22	NWB-1517	08/14/2020	9040C
08/14/2020	Water	L3690-03	pH	Cool 4 deg C	PSEG01	K22	NWB-1516	08/14/2020	9040C

Date/Time 08-14-2020 13:50
 Received by: JB (wet)
 Relinquished by: JB (Chen)

Date/Time 08-14-2020 15:00
 Received by: JB (wet)
 Relinquished by: JB (Chen)