

LB110540


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43965



A= 0.0000e+000

B= 2.2935e-004

C= -3.1313e-002

Rho= 0.9998454

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.002	0.002	145	0.000	145				
0.2	0.200	0.156	-0.044	817	0.0 %	817				
2.5	2.500	2.578	0.078	11377	0.0 %	11377				
5.0	5.000	5.029	0.029	22062	0.0 %	22062				
7.5	7.500	7.383	-0.117	32327	0.0 %	32327				
10.0	10.000	10.052	0.052	43965	0.0 %	43965				

LB110540 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 Aug 2020 14:22:28

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	145	-	0.00007471B	PPB		14 Aug 2020 14:24:33	S	Std
0.2 - 1	50.2	817	-	0.20007471B	PPB		14 Aug 2020 14:26:50	S	Std
2.5 - 1	52.5	11377	-	2.50007471B	PPB		14 Aug 2020 14:29:07	S	Std
5.0 - 1	55	22062	-	5.00007471B	PPB		14 Aug 2020 14:31:24	S	Std
7.5 - 1	57.5	32327	-	7.50007471B	PPB		14 Aug 2020 14:33:40	S	Std
10.0 - 1	510	43965	-	10.00007471B	PPB		14 Aug 2020 14:38:39	S	Std
ICV69 - 1	ICV69	18483	4.2078	-7471B	PPB		14 Aug 2020 14:41:54	U	SMPL
ICB69 - 1	ICB69	-42	-0.0409	-7471B	PPB		14 Aug 2020 14:44:09	U	SMPL
CCV22 - 1	CCV22	22346	5.0938	-7471B	PPB		14 Aug 2020 14:46:26	U	SMPL
CCB22 - 1	CCB22	-90	-0.0520	-7471B	PPB		14 Aug 2020 14:48:41	U	SMPL
CRA - 1	CRA	902	0.1756	-7471B	PPB		14 Aug 2020 14:50:58	U	SMPL
HighStd - 1	HighStd	43946	10.0479	-7471B	PPB		14 Aug 2020 14:53:14	U	SMPL
ChkStd - 1	ChkStd	32520	7.4273	-7471B	PPB		14 Aug 2020 14:55:30	U	SMPL
PB130986BL - 1	PBS	-225	-0.0829	-7471B	PPB		14 Aug 2020 14:57:48	U	SMPL
PB130986BS - 1	LCSS	16158	3.6746	-7471B	PPB		14 Aug 2020 15:00:05	U	SMPL
L3502-06 - 1	OK-01-081320	1809	0.3836	-7471B	PPB		14 Aug 2020 15:07:14	U	SMPL
L3502-06DUP - 1	OK-01-081320DUP	1621	0.3405	-7471B	PPB		14 Aug 2020 15:09:30	U	SMPL
L3502-06MS - 1	OK-01-081320MS	18779	4.2757	-7471B	PPB		14 Aug 2020 15:11:47	U	SMPL
L3502-06MSD - 1	OK-01-081320MSD	21284	4.8503	-7471B	PPB		14 Aug 2020 15:14:04	U	SMPL
L3681-02 - 1	TP-6-4-9FT	40552	9.2694	-7471B	PPB		14 Aug 2020 15:16:21	U	SMPL
CCV23 - 1	CCV23	22088	5.0347	-7471B	PPB		14 Aug 2020 15:18:37	U	SMPL
CCB23 - 1	CCB23	-89	-0.0517	-7471B	PPB		14 Aug 2020 15:20:55	U	SMPL
L3502-06LX5 - 1		372	0.0540	-7471B	PPB		14 Aug 2020 15:28:08	U	SMPL
L3502-06A - 1		20696	4.7154	-7471B	PPB		14 Aug 2020 15:30:23	U	SMPL
L3681-05 - 1	TP-2	218107	49.9924	-7471B	PPB		14 Aug 2020 15:37:42	U	SMPL
L3681-05DLX10 - 1		21674	4.9397	-7471B	PPB		14 Aug 2020 15:39:58	U	SMPL
CCV24 - 1	CCV24	22099	5.0372	-7471B	PPB		14 Aug 2020 15:42:27	U	SMPL
CCB24 - 1	CCB24	-136	-0.0625	-7471B	PPB		14 Aug 2020 15:44:46	U	SMPL