


LB110557

7470A

INSTRUMENT ID: CV2



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.080	0.080	-42	0.000	-42				
0.2	0.200	0.236	0.036	419	0.0 %	419				
2.5	2.500	2.452	-0.048	6931	0.0 %	6931				
5.0	5.000	4.837	-0.163	13943	0.0 %	13943				
7.5	7.500	7.528	0.028	21854	0.0 %	21854				
10.0	10.000	10.068	0.068	29321	0.0 %	29321				

LB110557 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Aug 2020 10:14:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-42	-	0.0000	7470A	PPB	17 Aug 2020 10:19:17	S	Std
0.2 - 1	50.2	419	-	0.2000	7470A	PPB	17 Aug 2020 10:21:25	S	Std
2.5 - 1	52.5	6931	-	2.5000	7470A	PPB	17 Aug 2020 10:23:32	S	Std
5.0 - 1	55	13943	-	5.0000	7470A	PPB	17 Aug 2020 10:25:40	S	Std
7.5 - 1	57.5	21854	-	7.5000	7470A	PPB	17 Aug 2020 10:27:47	S	Std
10.0 - 1	510	29321	-	10.0000	7470A	PPB	17 Aug 2020 10:29:54	S	Std
ICV70 - 1	ICV70	11783	4.1020	-7470A	PPB	PPB	17 Aug 2020 10:32:59	U	SMPL
ICB70 - 1	ICB70	-190	0.0292	-7470A	PPB	PPB	17 Aug 2020 10:35:04	U	SMPL
CCV25 - 1	CCV25	15194	5.2623	-7470A	PPB	PPB	17 Aug 2020 10:37:12	U	SMPL
CCB25 - 1	CCB25	-208	0.0231	-7470A	PPB	PPB	17 Aug 2020 10:39:18	U	SMPL
CRA - 1	CRA	426	0.2387	-7470A	PPB	PPB	17 Aug 2020 10:41:25	U	SMPL
HighStd - 1	HighStd	30104	10.3342	-7470A	PPB	PPB	17 Aug 2020 10:43:30	U	SMPL
ChkStd - 1	ChkStd	21221	7.3125	-7470A	PPB	PPB	17 Aug 2020 10:45:37	U	SMPL
PB130985BL - 1	PBW	-289	-0.0045	-7470A	PPB	PPB	17 Aug 2020 10:47:46	U	SMPL
PB130985BS - 1	LCSW	10301	3.5979	-7470A	PPB	PPB	17 Aug 2020 10:49:55	U	SMPL
L3502-10 - 1	OK-01-081320	-190	0.0292	-7470A	PPB	PPB	17 Aug 2020 10:52:02	U	SMPL
L3502-10DUP - 1	OK-01-081320DUP	-42	0.0795	-7470A	PPB	PPB	17 Aug 2020 10:54:10	U	SMPL
L3502-10MS - 1	OK-01-081320MS	9679	3.3863	-7470A	PPB	PPB	17 Aug 2020 10:56:17	U	SMPL
L3502-10MSD - 1	OK-01-081320MSD	11316	3.9432	-7470A	PPB	PPB	17 Aug 2020 10:58:25	U	SMPL
L3681-03 - 1	TP-6-4-9FT	-142	0.0455	-7470A	PPB	PPB	17 Aug 2020 11:00:32	U	SMPL
CCV26 - 1	CCV26	14462	5.0133	-7470A	PPB	PPB	17 Aug 2020 11:11:27	U	SMPL
CCB26 - 1	CCB26	-218	0.0197	-7470A	PPB	PPB	17 Aug 2020 11:13:33	U	SMPL
L3681-06 - 1	TP-2	-14	0.0891	-7470A	PPB	PPB	17 Aug 2020 11:15:41	U	SMPL
L3682-06 - 1	20200513-2801-A	17	0.0996	-7470A	PPB	PPB	17 Aug 2020 11:17:47	U	SMPL
L3682-07 - 1	20200514-2801-B	7	0.0962	-7470A	PPB	PPB	17 Aug 2020 11:19:53	U	SMPL
L3682-08 - 1	20200515-2801-C	7	0.0962	-7470A	PPB	PPB	17 Aug 2020 11:23:16	U	SMPL
L3682-09 - 1	20200516-2801-D	9	0.0969	-7470A	PPB	PPB	17 Aug 2020 11:25:23	U	SMPL
L3682-10 - 1	20200517-2801-E	37	0.1064	-7470A	PPB	PPB	17 Aug 2020 11:27:30	U	SMPL
PB130985TB - 1	PB130953TB	4	0.0952	-7470A	PPB	PPB	17 Aug 2020 11:29:37	U	SMPL
L3502-10LX5 - 1		-12	0.0898	-7470A	PPB	PPB	17 Aug 2020 11:31:45	U	SMPL
L3502-10A - 1		11419	3.9782	-7470A	PPB	PPB	17 Aug 2020 11:33:53	U	SMPL
CCV27 - 1	CCV27	15168	5.2535	-7470A	PPB	PPB	17 Aug 2020 11:38:49	U	SMPL
CCB27 - 1	CCB27	-45	0.0785	-7470A	PPB	PPB	17 Aug 2020 11:42:29	U	SMPL