


LB110566

ISM02.4

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.3090e-004

C= -4.1635e-002

Rho= 0.9999170

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.024	0.024	286	0.000	286					—
0.05	0.050					159					4
0.20	0.200	0.207	0.007	1077	0.0 %	1077					-2
2.50	2.500	2.462	-0.038	10843	0.0 %	10843					1
5.00	5.000	5.040	0.040	22006	0.0 %	22006					-1
7.50	7.500	7.415	-0.085	32294	0.0 %	32294					1
10.0	10.000	10.052	0.052	43714	0.0 %	43714					

LB110566
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	286	-		0 ISM02.4	PPB	8/17/2020 14:01	Std
	0.2 S0.2	1077	-		0.2 ISM02.4	PPB	8/17/2020 14:04	Std
	2.5 S2.5	10843	-		2.5 ISM02.4	PPB	8/17/2020 14:08	Std
	5 S5.0	22006	-		5 ISM02.4	PPB	8/17/2020 14:10	Std
	7.5 S7.5	32294	-		7.5 ISM02.4	PPB	8/17/2020 14:12	Std
	10 S10.0	43714	-		10 ISM02.4	PPB	8/17/2020 14:14	Std
ICV	ICV	16569	3.7841 -		ISM02.4	PPB	8/17/2020 14:17	SMPL
ICB	ICB	104	-0.0176 -		ISM02.4	PPB	8/17/2020 14:19	SMPL
CCV028	CCV028	22380	5.1259 -		ISM02.4	PPB	8/17/2020 14:22	SMPL
CCB028	CCB028	86	-0.0218 -		ISM02.4	PPB	8/17/2020 14:24	SMPL
PB131010BL	PBW010	177	-0.0008 -		ISM02.4	PPB	8/17/2020 14:26	SMPL
L3686-07	MX2N01	28478	6.5339 -		ISM02.4	PPB	8/17/2020 14:34	SMPL
L3686-08	MX2N01D	27696	6.3534 -		ISM02.4	PPB	8/17/2020 14:36	SMPL
L3686-09	MX2N01S	32832	7.5393 -		ISM02.4	PPB	8/17/2020 14:41	SMPL
CCV029	CCV029	21985	5.0347 -		ISM02.4	PPB	8/17/2020 14:43	SMPL
CCB029	CCB029	167	-0.0031 -		ISM02.4	PPB	8/17/2020 14:48	SMPL