



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB110568

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	08/17/2020	09:35
2	L3690-04		1	8	100.00	>212.0	>212.0	08/17/2020	10:05
3	L3690-04DUP		1	8	100.00	>212.0	>212.0	08/17/2020	10:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB110568

WorkList Name : fp81720 WorkList ID : 141520 Department : Wet-Chemistry Date : 08-17-2020 09:02:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/28/2020	Water	L3690-04 B	Flash Point	Cool 4 deg C	PSEG01	K22	NWB-1518-1517-1516	08/14/2020	1010A

Date/Time 08/17/2020 09:10
 Received by: KS (watcher)
 Relinquished by: KS (watcher)

Date/Time 08/17/2020 10:45
 Received by: KS (watcher)
 Relinquished by: KS (watcher)