



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110572

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	3.99	08/17/2020	16:13
2	CAL2	1	Water	NA	NA	21.3	7.00	08/17/2020	16:14
3	CAL3	1	Water	NA	NA	21.3	10.03	08/17/2020	16:15
4	ICV	1	Water	NA	NA	21.4	7.00	08/17/2020	16:19
5	CCV1	1	Water	NA	NA	21.3	2.01	08/17/2020	16:20
6	L3712-15	1	Water	NA	NA	24.1	6.68	08/17/2020	16:22
7	L3712-15DUP	1	Water	NA	NA	24.2	6.69	08/17/2020	16:23
8	CCV2	1	Water	NA	NA	21.3	12.02	08/17/2020	16:25

VB 110572

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3712

WorkList ID : 141541

Department : Wet-Chemistry

Date : 08-17-2020 16:01:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/17/2020	Water	L3712-15	pH -F	Cool 4 deg C	PSEG01	K41	VB-25860	08/17/2020	9040C

Date/Time 08-17-2020 16:10
 Received by: J2 (wet)
 Relinquished by: J2 (wet) (only)

Date/Time 08-17-2020 17:10
 Received by: J2 (wet)
 Relinquished by: J2 (only)