


LB110575

7470A

INSTRUMENT ID: CV2



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.096	-0.096	-34	0.000	-34				
0.2	0.200	0.183	-0.017	918	0.0 %	918				
2.5	2.500	2.602	0.102	9189	0.0 %	9189				
5.0	5.000	4.977	-0.023	17310	0.0 %	17310				
7.5	7.500	7.739	0.239	26755	0.0 %	26755				
10.0	10.000	9.795	-0.205	33785	0.0 %	33785				

LB110575 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 18 Aug 2020 09:29:13					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-34	-	0.00007470A	PPB		18 Aug 2020 09:33:21	S	Std
0.2 - 1	50.2	918	-	0.20007470A	PPB		18 Aug 2020 09:35:28	S	Std
2.5 - 1	52.5	9189	-	2.50007470A	PPB		18 Aug 2020 09:37:37	S	Std
5.0 - 1	55	17310	-	5.00007470A	PPB		18 Aug 2020 09:39:44	S	Std
7.5 - 1	57.5	26755	-	7.50007470A	PPB		18 Aug 2020 09:44:16	S	Std
10.0 - 1	510	33785	-	10.00007470A	PPB		18 Aug 2020 09:48:50	S	Std
ICV71 - 1	ICV71	13099	3.7453	-7470A	PPB		18 Aug 2020 09:51:36	U	SMPL
ICB71 - 1	ICB71	25	-0.0784	-7470A	PPB		18 Aug 2020 09:53:41	U	SMPL
CCV34 - 1	CCV34	18519	5.3304	-7470A	PPB		18 Aug 2020 09:55:49	U	SMPL
CCB34 - 1	CCB34	6	-0.0840	-7470A	PPB		18 Aug 2020 09:57:54	U	SMPL
CRA - 1	CRA	833	0.1579	-7470A	PPB		18 Aug 2020 10:00:02	U	SMPL
HighStd - 1	HighStd	34226	9.9242	-7470A	PPB		18 Aug 2020 10:02:07	U	SMPL
ChkStd - 1	ChkStd	24038	6.9445	-7470A	PPB		18 Aug 2020 10:04:13	U	SMPL
PB130926BL - 1	PBW	145	-0.0433	-7470A	PPB		18 Aug 2020 10:08:50	U	SMPL
PB130926BS - 1	LCSW	12310	3.5145	-7470A	PPB		18 Aug 2020 10:13:18	U	SMPL
L3645-03 - 1	GW-3	1038	0.2179	-7470A	PPB		18 Aug 2020 10:15:25	U	SMPL
L3645-04 - 1	GW-3	-165	-0.1340	-7470A	PPB		18 Aug 2020 10:17:33	U	SMPL
L3680-01 - 1	DRUM-1-8	27	-0.0778	-7470A	PPB		18 Aug 2020 10:19:42	U	SMPL
L3680-01DUP - 1	DRUM-1-8DUP	66	-0.0664	-7470A	PPB		18 Aug 2020 10:21:50	U	SMPL
L3680-01MS - 1	DRUM-1-8MS	11898	3.3940	-7470A	PPB		18 Aug 2020 10:23:56	U	SMPL
CCV35 - 1	CCV35	18240	5.2488	-7470A	PPB		18 Aug 2020 10:26:02	U	SMPL
CCB35 - 1	CCB35	23	-0.0790	-7470A	PPB		18 Aug 2020 10:28:08	U	SMPL
L3680-01MSD - 1	DRUM-1-8MSD	11827	3.3733	-7470A	PPB		18 Aug 2020 10:30:15	U	SMPL
L3719-07 - 1	20200508-FIELD-BLANK	-202	-0.1448	-7470A	PPB		18 Aug 2020 10:32:21	U	SMPL
PB131027BL - 1	PBW	-11	-0.0889	-7470A	PPB		18 Aug 2020 10:34:27	U	SMPL
L3500-08 - 1	CL-02-081420	-155	-0.1311	-7470A	PPB		18 Aug 2020 10:38:40	U	SMPL
L3500-08DUP - 1	CL-02-081420DUP	-30	-0.0945	-7470A	PPB		18 Aug 2020 10:40:48	U	SMPL
L3500-08MS - 1	CL-02-081420MS	12829	3.6663	-7470A	PPB		18 Aug 2020 10:46:14	U	SMPL
L3500-08MSD - 1	CL-02-081420MSD	12785	3.6534	-7470A	PPB		18 Aug 2020 10:48:21	U	SMPL
L3508-22 - 1	TR-08-081420	-152	-0.1302	-7470A	PPB		18 Aug 2020 10:50:29	U	SMPL
CCV36 - 1	CCV36	18844	5.4255	-7470A	PPB		18 Aug 2020 10:52:36	U	SMPL
CCB36 - 1	CCB36	78	-0.0629	-7470A	PPB		18 Aug 2020 10:54:42	U	SMPL
PB131027BS - 1	LCSW	12674	3.6210	-7470A	PPB		18 Aug 2020 10:56:50	U	SMPL
LB110575TB - 1	PB130987TB	-185	-0.1398	-7470A	PPB		18 Aug 2020 10:58:56	U	SMPL
L3680-01LX5 - 1		-36	-0.0962	-7470A	PPB		18 Aug 2020 11:01:03	U	SMPL
L3680-01A - 1		13985	4.0044	-7470A	PPB		18 Aug 2020 11:03:10	U	SMPL
L3500-08LX5 - 1		-195	-0.1428	-7470A	PPB		18 Aug 2020 11:05:16	U	SMPL
L3500-08A - 1		13932	3.9889	-7470A	PPB		18 Aug 2020 11:07:24	U	SMPL
CCV37 - 1	CCV37	18133	5.2175	-7470A	PPB		18 Aug 2020 11:09:31	U	SMPL
CCB37 - 1	CCB37	-45	-0.0989	-7470A	PPB		18 Aug 2020 11:11:39	U	SMPL

LB131027A m3
8/18/20