



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB110577

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	08/18/2020	10:15
2	CAL2	1	Water	NA	NA	21.1	7.00	08/18/2020	10:16
3	CAL3	1	Water	NA	NA	21.1	10.03	08/17/2020	10:19
4	ICV	1	Water	NA	NA	21.1	6.99	08/18/2020	10:20
5	CCV1	1	Water	NA	NA	21.2	2.01	08/18/2020	10:22
6	L3710-06	1	Water	NA	NA	16.8	6.72	08/18/2020	10:24
7	L3710-06DUP	1	Water	NA	NA	16.9	6.73	08/18/2020	10:25
8	CCV2	1	Water	NA	NA	21.3	12.02	08/18/2020	10:26

283 110577

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L3696 w WorkList ID : 141568 Department : Wet-Chemistry Date : 08-18-2020 10:05:27

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/17/2020	Water	L3710-06	Corrosivity	Cool 4 deg C	PSEG01	K41	1-5-WATER-DRUM	08/17/2020	9040C

Date/Time 08/18/2020 10:10
Received by: JP (wet)
Relinquished by: JP (wet)

Date/Time 08/18/2020 11:30
Received by: JP (wet)
Relinquished by: JP (wet)