



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 8/19/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 08/18/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:50
Out Date: 08/19/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB110578

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
L3507-08	SU-02-081820	9	1.17	9.7	8.7	88.3	
L3507-09	SU-02-081820	10	1.15	9.73	8.76	88.7	
L3507-10	SU-02-081820-1	11	1.17	9.57	8.61	88.6	
L3507-11	SU-02-081820-2	12	1.16	9.63	9.19	94.8	
L3710-01	1-3-SOIL-DRUM	1	1.15	9.73	8.24	82.6	
L3710-02	1-5-OILY-DRUM	2	1.19	9.9	8.2	80.5	
L3710-04	5-9-OILY-DRUM	3	1.14	9.93	8.43	82.9	
L3710-09	1-3-SOIL-DRUM-EPH-2	4	1.15	9.73	8.24	82.6	
L3713-01	SIDEWALK-CONCRETE	5	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
L3714-01	WHITEHORSE-CONCRETE	6	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
L3716-01	PIPE-1	7	1.00	2.00	2.00	100.0	WIPE SAMPLE
L3716-02	PIPE-2	8	1.00	2.00	2.00	100.0	WIPE SAMPLE
L3724-01	T1-FIREWALL	13	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$