

Analysis Method: 9034
 Parameter: Reactive Sulfide
 Run Number: LB110587

ANALYST: ketankumar
 SUPERVISOR REVIEW BY: apatel
 Constant: 16000
 Normality1: 0.025
 Normality2: 0.025

Reagent/Standard	Lot/Log #
Sodium thiosulfate titrant 0.025N	WP84605
IODINE SOLUTION .025N 1L	W2670
Starch Solution, 4L	W2703

Seq	Lab ID	True Value (mg/l)	DF	Initial Weight (g)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB131056BL		1	5.02	50	2.00	0.00	1.88	1.88	0.12	0.00	0.00	08/18/2020	13:30
2	L3496-06		1	5.02	50	2.00	1.88	3.72	1.84	0.16	0.04	3.19	08/18/2020	13:32
3	L3496-06DUP		1	5.04	50	2.00	3.72	5.58	1.86	0.14	0.02	1.59	08/18/2020	13:35
4	L3500-08		1	5.05	50	2.00	5.58	7.40	1.82	0.18	0.06	4.75	08/18/2020	13:38
5	L3508-22		1	5.00	50	2.00	0.00	1.84	1.84	0.16	0.04	3.20	08/18/2020	13:40
6	L3710-01		1	5.02	50	2.00	1.84	3.64	1.80	0.20	0.08	6.37	08/18/2020	13:42
7	L3710-02		1	5.06	50	2.00	3.64	5.48	1.84	0.16	0.04	3.16	08/18/2020	13:45
8	L3710-04		1	5.03	50	2.00	5.48	7.28	1.80	0.20	0.08	6.36	08/18/2020	13:48

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume