



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB110593

Slope : 97.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.00	08/18/2020	17:30
2	CAL2	1	Water	NA	NA	21.1	7.00	08/18/2020	17:32
3	CAL3	1	Water	NA	NA	21.2	10.03	08/18/2020	17:33
4	ICV	1	Water	NA	NA	21.2	6.98	08/18/2020	17:35
5	CCV1	1	Water	NA	NA	21.2	2.02	08/18/2020	17:36
6	L3499-04	1	Solid	20.03	20	23.8	9.13	08/19/2020	17:39
7	L3507-12	1	Solid	20.02	20	24.0	8.77	08/18/2020	17:40
8	L3507-12DUP	1	Solid	20.03	20	23.9	8.78	08/18/2020	17:41
9	CCV2	1	Water	NA	NA	21.3	12.02	08/18/2020	17:43

YB110593

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSION L3681 WorkList ID : 141585 Department : Wet-Chemistry Date : 08-18-2020 13:20:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/18/2020	Solid	L3499-04	Corrosivity	Cool 4 deg C	PSEG05	K43	HD-01-081820	08/18/2020	9045D
08/18/2020	Solid	L3507-12	Corrosivity	Cool 4 deg C	PSEG05		SU-02-081820	08/18/2020	9045D

Date/Time 08-18-2020 16:56
Received by: JD (wet-chem)
Relinquished by: JA

Date/Time 08-18-2020 18:10
Received by: JD
Relinquished by: J. A. L. W. E. T. C. H. E. M.