


LB110595

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.023	0.023	52	0.000	52				
0.2	0.200	0.223	0.023	711	0.0 %	711				
2.5	2.500	2.487	-0.013	8169	0.0 %	8169				
5.0	5.000	4.895	-0.105	16105	0.0 %	16105				
7.5	7.500	7.570	0.070	24920	0.0 %	24920				
10.0	10.000	10.001	0.001	32929	0.0 %	32929				

LB110595 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 19 Aug 2020 09:18:01					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	52	-	0.0000	7470A	PPB	19 Aug 2020 10:13:41	S	Std
0.2 - 1	50.2	711	-	0.2000	7470A	PPB	19 Aug 2020 10:15:49	S	Std
2.5 - 1	52.5	8169	-	2.5000	7470A	PPB	19 Aug 2020 10:17:57	S	Std
5.0 - 1	55	16105	-	5.0000	7470A	PPB	19 Aug 2020 10:20:05	S	Std
7.5 - 1	59.5	24920	-	7.5000	7470A	PPB	19 Aug 2020 10:24:39	S	Std
10.0 - 1	510	32929	-	10.0000	7470A	PPB	19 Aug 2020 10:29:26	S	Std
ICV73 - 1	ICV73	13411	4.0776	-	7470A	PPB	19 Aug 2020 10:32:02	U	SMPL
ICB73 - 1	ICB73	-155	-0.0393	-	7470A	PPB	19 Aug 2020 10:34:08	U	SMPL
CCV42 - 1	CCV42	17682	5.3737	-	7470A	PPB	19 Aug 2020 10:36:14	U	SMPL
CCB42 - 1	CCB42	-154	-0.0390	-	7470A	PPB	19 Aug 2020 10:38:20	U	SMPL
CRA - 1	CRA	705	0.2216	-	7470A	PPB	19 Aug 2020 10:40:26	U	SMPL
HighStd - 1	HighStd	33147	10.0670	-	7470A	PPB	19 Aug 2020 10:42:32	U	SMPL
ChkStd - 1	ChkStd	24988	7.5909	-	7470A	PPB	19 Aug 2020 10:44:39	U	SMPL
PB131062BL - 1	PBW	59	0.0256	-	7470A	PPB	19 Aug 2020 10:49:11	U	SMPL
PB131062BS - 1	LCSW	11922	3.6257	-	7470A	PPB	19 Aug 2020 10:51:17	U	SMPL
L3719-08 - 1	SPLP-AMENDED-DISCRETE-004	-83	-0.0175	-	7470A	PPB	19 Aug 2020 10:53:25	U	SMPL
L3719-09 - 1	SPLP-AMENDED-DISCRETE-005	39	0.0195	-	7470A	PPB	19 Aug 2020 10:55:32	U	SMPL
L3719-10 - 1	SPLP-AMENDED-COMP-A	46	0.0217	-	7470A	PPB	19 Aug 2020 10:57:39	U	SMPL
L3719-10DUP - 1	SPLP-AMENDED-COMP-ADUP	54	0.0241	-	7470A	PPB	19 Aug 2020 10:59:47	U	SMPL
L3719-10MS - 1	SPLP-AMENDED-COMP-AMS	14087	4.2827	-	7470A	PPB	19 Aug 2020 11:01:53	U	SMPL
CCV43 - 1	CCV43	16472	5.0065	-	7470A	PPB	19 Aug 2020 11:03:59	U	SMPL
CCB43 - 1	CCB43	-131	-0.0321	-	7470A	PPB	19 Aug 2020 11:06:06	U	SMPL
L3719-10MSD - 1	SPLP-AMENDED-COMP-AMSD	13615	4.1395	-	7470A	PPB	19 Aug 2020 11:13:14	U	SMPL
PB131063BL - 1	PBW	-124	-0.0299	-	7470A	PPB	19 Aug 2020 11:15:20	U	SMPL
PB131063BS - 1	LCSW	11570	3.5189	-	7470A	PPB	19 Aug 2020 11:17:26	U	SMPL
L3496-06 - 1	AU-06-081720	-109	-0.0254	-	7470A	PPB	19 Aug 2020 11:19:32	U	SMPL
L3496-06DUP - 1	AU-06-081720DUP	13	0.0116	-	7470A	PPB	19 Aug 2020 11:21:39	U	SMPL
L3496-06MS - 1	AU-06-081720MS	13221	4.0199	-	7470A	PPB	19 Aug 2020 11:23:46	U	SMPL
L3496-06MSD - 1	AU-06-081720MSD	11933	3.6291	-	7470A	PPB	19 Aug 2020 11:25:53	U	SMPL
L3689-03 - 1	TP-1-4-8FT	-118	-0.0281	-	7470A	PPB	19 Aug 2020 11:28:01	U	SMPL
L3689-06 - 1	TP-5	18	0.0132	-	7470A	PPB	19 Aug 2020 11:30:09	U	SMPL
L3689-09 - 1	TP-4	33	0.0177	-	7470A	PPB	19 Aug 2020 11:32:15	U	SMPL
CCV44 - 1	CCV44	16372	4.9762	-	7470A	PPB	19 Aug 2020 11:34:21	U	SMPL
CCB44 - 1	CCB44	-116	-0.0275	-	7470A	PPB	19 Aug 2020 11:36:28	U	SMPL
L3689-12 - 1	TP-3	30	0.0168	-	7470A	PPB	19 Aug 2020 11:38:35	U	SMPL
L3712-01 - 1	CHRT-27137	48	0.0223	-	7470A	PPB	19 Aug 2020 11:40:41	U	SMPL
L3712-03 - 1	CHRT-26572	67	0.0280	-	7470A	PPB	19 Aug 2020 11:42:47	U	SMPL
L3712-05 - 1	VB-25203	79	0.0317	-	7470A	PPB	19 Aug 2020 11:44:54	U	SMPL
L3712-07 - 1	VB-27207	66	0.0277	-	7470A	PPB	19 Aug 2020 11:47:01	U	SMPL
L3712-09 - 1	V-23615	70	0.0289	-	7470A	PPB	19 Aug 2020 11:49:08	U	SMPL
L3712-11 - 1	CHRT-25433	70	0.0289	-	7470A	PPB	19 Aug 2020 11:51:15	U	SMPL
L3712-13 - 1	CHRT-26982	72	0.0295	-	7470A	PPB	19 Aug 2020 11:53:22	U	SMPL
PB131064BL - 1	PBW	71	0.0292	-	7470A	PPB	19 Aug 2020 11:55:30	U	SMPL
PB131064BS - 1	LCSW	12753	3.8779	-	7470A	PPB	19 Aug 2020 11:57:36	U	SMPL
CCV45 - 1	CCV45	16521	5.0214	-	7470A	PPB	19 Aug 2020 11:59:43	U	SMPL
CCB45 - 1	CCB45	-122	-0.0293	-	7470A	PPB	19 Aug 2020 12:01:49	U	SMPL
L3680-02 - 1	DRUM-1-8	138	0.0496	-	7470A	PPB	19 Aug 2020 12:03:57	U	SMPL
L3712-15 - 1	VB-25860	31	0.0171	-	7470A	PPB	19 Aug 2020 12:06:03	U	SMPL
L3712-15DUP - 1	VB-25860DUP	63	0.0268	-	7470A	PPB	19 Aug 2020 12:08:09	U	SMPL
L3712-15MS - 1	VB-25860MS	12250	3.7253	-	7470A	PPB	19 Aug 2020 12:10:15	U	SMPL
L3712-15MSD - 1	VB-25860MSD	12465	3.7905	-	7470A	PPB	19 Aug 2020 12:12:22	U	SMPL
L3715-04 - 1	226	13	0.0116	-	7470A	PPB	19 Aug 2020 12:14:29	U	SMPL
PB131065BL - 1	PBW	-66	-0.0123	-	7470A	PPB	19 Aug 2020 12:16:37	U	SMPL
PB131065BS - 1	LCSW	11991	3.6467	-	7470A	PPB	19 Aug 2020 12:18:44	U	SMPL
L3710-06 - 1	1-5-WATER-DRUM	17586	5.3446	-	7470A	PPB	19 Aug 2020 12:20:53	U	SMPL
L3721-01 - 1	I-OC-6	-293	-0.0812	-	7470A	PPB	19 Aug 2020 12:22:59	U	SMPL
CCV46 - 1	CCV46	16246	4.9379	-	7470A	PPB	19 Aug 2020 12:25:09	U	SMPL
CCB46 - 1	CCB46	-26	-0.0002	-	7470A	PPB	19 Aug 2020 12:27:16	U	SMPL
L3721-05 - 1	I-R-6	949	0.2957	-	7470A	PPB	19 Aug 2020 12:29:22	U	SMPL
L3721-06 - 1	E-R-6	286	0.0945	-	7470A	PPB	19 Aug 2020 12:31:28	U	SMPL
L3721-06DUP - 1	E-R-6DUP	252	0.0842	-	7470A	PPB	19 Aug 2020 12:33:36	U	SMPL
L3721-06MS - 1	E-R-6MS	13259	4.0315	-	7470A	PPB	19 Aug 2020 12:35:42	U	SMPL
L3721-06MSD - 1	E-R-6MSD	13201	4.0139	-	7470A	PPB	19 Aug 2020 12:37:49	U	SMPL
PB131062TB - 1	PB131030TB	-65	-0.0120	-	7470A	PPB	19 Aug 2020 12:39:57	U	SMPL
PB131063TB - 1	PB131031TB	69	0.0286	-	7470A	PPB	19 Aug 2020 12:42:05	U	SMPL
PB131064TB - 1	PB131032TB	80	0.0320	-	7470A	PPB	19 Aug 2020 12:44:13	U	SMPL
CCV47 - 1	CCV47	16045	4.8770	-	7470A	PPB	19 Aug 2020 12:46:21	U	SMPL
CCB47 - 1	CCB47	1	0.0080	-	7470A	PPB	19 Aug 2020 12:48:28	U	SMPL
L3719-10LX5 - 1		60	0.0259	-	7470A	PPB	19 Aug 2020 12:50:34	U	SMPL
L3719-10A - 1		13891	4.2233	-	7470A	PPB	19 Aug 2020 12:52:42	U	SMPL
L3496-06LX5 - 1		-10	0.0047	-	7470A	PPB	19 Aug 2020 12:54:49	U	SMPL
L3496-06A - 1		13583	4.1298	-	7470A	PPB	19 Aug 2020 12:56:56	U	SMPL
L3712-15LX5 - 1		-31	-0.0017	-	7470A	PPB	19 Aug 2020 12:59:03	U	SMPL
L3712-15A - 1		13347	4.0582	-	7470A	PPB	19 Aug 2020 13:01:10	U	SMPL
L3721-06LX5 - 1		126	0.0459	-	7470A	PPB	19 Aug 2020 13:06:54	U	SMPL
L3721-06A - 1		13666	4.1550	-	7470A	PPB	19 Aug 2020 13:09:01	U	SMPL
CCV48 - 1	CCV48	15975	4.8557	-	7470A	PPB	19 Aug 2020 13:11:08	U	SMPL
CCB48 - 1	CCB48	-49	-0.0072	-	7470A	PPB	19 Aug 2020 13:13:18	U	SMPL