



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB110601

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP86375
Turbidity Calibration std, 20NTU	WP86371
Turbidity Calibration std, 40NTU	WP86370
Turbidity Calibration std, 1NTU	WP86374
Turbidity Calibration std, 5NTU	WP86373
Turbidity Calibration - CCV std, 10 NTU	WP86372
10 NTU Standard 500 ml	W2489

Intercept: 0.3346

Slope: 0.9589

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.254	08/19/2020	12:20
2	CAL2	1	1	1.250	08/19/2020	12:22
3	CAL3	5	1	5.200	08/19/2020	12:24
4	CAL4	10	1	9.880	08/19/2020	12:26
5	CAL5	20	1	19.700	08/19/2020	12:29
6	CAL6	40	1	38.600	08/19/2020	12:32



Analytical Summary Report

Analysis Method: SM2130 B
Parameter: Turbidity
Run Number: LB110601

ANALYST: Alexander
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.300	10.393	08/19/2020	12:36
2	ICB		1	0.248	-0.090	08/19/2020	12:38
3	CCV1	10	1	9.900	9.975	08/19/2020	12:41
4	CCB1		1	0.242	-0.097	08/19/2020	12:43
5	LB110601BL		1	0.233	-0.106	08/19/2020	12:45
6	L3739-01		1	13.100	13.313	08/19/2020	12:47
7	L3739-01DUP		1	12.800	13.000	08/19/2020	12:49
8	CCV2	10	1	9.980	10.059	08/19/2020	12:51
9	CCB2		1	0.256	-0.082	08/19/2020	12:53

LB110601

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Turbidity-81920

WorkList ID : 141622

Department : Wet-Chemistry

Date : 08-19-2020 11:50:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/20/2020	Water	L3739-01	Turbidity	Cool 4 deg C	LIRO01	K61	UDS-6-16A	08/18/2020	SM2130 B

Date/Time 8/19/2020 11:55

Received by: Ann(Wet-chem)

Relinquished by: Jo (wet-chem)

Date/Time 8/19/2020 12:55

Received by: Jo (wet-chem)

Relinquished by: Ann(Wet-chem)