



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB110604

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	08/19/2020	09:35
2	L3710-06		1	8	100.00	>212.0	>212.0	08/19/2020	10:10
3	L3710-06DUP		1	8	100.00	>212.0	>212.0	08/19/2020	10:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp81820 WorkList ID : 141558 Department : Wet-Chemistry Date : 08-18-2020 08:57:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/24/2020	Water	L3710-06	Flash Point	Cool 4 deg C	PSEG01	K41	1-5-WATER-DRUM	08/17/2020	1010A

Date/Time 08/19/2020 09:00
 Received by: KS (wekchen)
 Relinquished by: KS (wekchen)

Date/Time 08/19/2020 10:45
 Received by: KS (wekchen)
 Relinquished by: KS (wekchen)