



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB110607

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	08/19/2020	17:15
2	CAL2	1	Water	NA	NA	21.1	7.00	08/19/2020	17:16
3	CAL3	1	Water	NA	NA	21.1	10.03	08/19/2020	17:19
4	ICV	1	Water	NA	NA	21.2	7.00	08/19/2020	17:20
5	CCV1	1	Water	NA	NA	21.2	2.02	08/19/2020	17:22
6	L3502-15	1	Solid	20.03	20	23.8	9.26	08/19/2020	17:25
7	L3504-12	1	Solid	20.04	20	23.9	9.17	08/19/2020	17:26
8	L3742-04	1	Solid	20.02	20	23.9	6.44	08/19/2020	17:29
9	L3742-08	1	Solid	20.03	20	23.9	6.60	08/19/2020	17:30
10	L3742-08DUP	1	Solid	20.04	20	23.8	6.61	08/19/2020	17:32
11	CCV2	1	Water	NA	NA	21.3	12.02	08/19/2020	17:35

LB110601

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L3732 WorkList ID : 141627 Department : Wet-Chemistry Date : 08-19-2020 13:35:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/19/2020	Solid	L3502-15	Corrosivity	Cool 4 deg C	PSEG05		OK-02-081920	08/19/2020	9045D
08/19/2020	Solid	L3504-12	Corrosivity	Cool 4 deg C	PSEG05		EO-01-081920	08/19/2020	9045D
08/19/2020	Solid	L3742-04	Corrosivity	Cool 4 deg C	PSEG01	K43	TP-10-S	08/19/2020	9045D
08/19/2020	Solid	L3742-08	Corrosivity	Cool 4 deg C	PSEG01	K43	TP-10-D	08/19/2020	9045D

Date/Time 08-19-2020 16:05
 Received by: In wet Chem)
 Relinquished by: al smi

Date/Time 08-19-2020 17:40
 Received by: CP smi
 Relinquished by: Jo wet Chem)