


LB110621

7471B

INSTRUMENT ID: C11

Linear



A= 0.0000e+000

B= 1.9976e-004

C= -7.2532e-003

Rho= 0.9999310

Accept = Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.006	0.006	67	0.000	67				
0.2	0.200	0.186	-0.014	967	0.0 %	967				
2.5	2.500	2.491	-0.009	12507	0.0 %	12507				
5.0	5.000	4.988	-0.012	25004	0.0 %	25004				
7.5	7.500	7.587	0.087	38015	0.0 %	38015				
10.0	10.000	9.943	-0.057	49809	0.0 %	49809				

LB110621 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 Aug 2020 13:31:05

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	67	-	0.0000	7471B	PPB	20 Aug 2020 13:32:59	S	Std
0.2 - 1	50.2	967	-	0.2000	7471B	PPB	20 Aug 2020 13:35:16	S	Std
2.5 - 1	52.5	12507	-	2.5000	7471B	PPB	20 Aug 2020 13:37:34	S	Std
5.0 - 1	55	25004	-	5.0000	7471B	PPB	20 Aug 2020 13:39:51	S	Std
7.5 - 1	57.5	38015	-	7.5000	7471B	PPB	20 Aug 2020 13:44:41	S	Std
10.0 - 1	510	49809	-	10.0000	7471B	PPB	20 Aug 2020 13:49:34	S	Std
ICV76 - 1	ICV76	20336	4.0551	-	7471B	PPB	20 Aug 2020 13:52:16	U	SMPL
ICB76 - 1	ICB76	-131	-0.0334	-	7471B	PPB	20 Aug 2020 13:54:32	U	SMPL
CCV55 - 1	CCV55	26041	5.1947	-	7471B	PPB	20 Aug 2020 13:56:48	U	SMPL
CCB55 - 1	CCB55	-162	-0.0396	-	7471B	PPB	20 Aug 2020 13:59:03	U	SMPL
CRA - 1	CRA	951	0.1827	-	7471B	PPB	20 Aug 2020 14:01:20	U	SMPL
HighStd - 1	HighStd	49925	9.9658	-	7471B	PPB	20 Aug 2020 14:03:35	U	SMPL
ChkStd - 1	ChkStd	36773	7.3385	-	7471B	PPB	20 Aug 2020 14:05:51	U	SMPL
PB131106BL - 1	PBS	-208	-0.0488	-	7471B	PPB	20 Aug 2020 14:08:09	U	SMPL
PB131106BS - 1	LCSS	18235	3.6354	-	7471B	PPB	20 Aug 2020 14:10:26	U	SMPL
L3502-11 - 1	OK-02-081920	8312	1.6532	-	7471B	PPB	20 Aug 2020 14:12:42	U	SMPL
L3504-07 - 1	EO-01-081920	6417	1.2746	-	7471B	PPB	20 Aug 2020 14:15:00	U	SMPL
L3720-03 - 1	MP-081920-B10-0-2-COMP	32015	6.3881	-	7471B	PPB	20 Aug 2020 14:17:18	U	SMPL
L3720-05 - 1	MP-081920-B11-0-2-COMP	86901	17.3521	-	7471B	PPB	20 Aug 2020 14:19:36	U	SMPL
L3720-07 - 1	MP-081920-BD-0-2-COMP	59143	11.8072	-	7471B	PPB	20 Aug 2020 14:21:53	U	SMPL
CCV56 - 1	CCV56	24857	4.9582	-	7471B	PPB	20 Aug 2020 14:24:11	U	SMPL
CCB56 - 1	CCB56	-156	-0.0384	-	7471B	PPB	20 Aug 2020 14:26:28	U	SMPL
L3720-09 - 1	MP-081920-B9-0-2-COMP	45959	9.1735	-	7471B	PPB	20 Aug 2020 14:28:45	U	SMPL
L3729-14 - 1	COMP-1-(0-1)	3426	0.6771	-	7471B	PPB	20 Aug 2020 14:31:00	U	SMPL
L3729-14DUP - 1	COMP-1-(0-1)DUP	3690	0.7299	-	7471B	PPB	20 Aug 2020 14:35:59	U	SMPL
L3729-14MS - 1	COMP-1-(0-1)MS	22253	4.4380	-	7471B	PPB	20 Aug 2020 14:38:14	U	SMPL
L3729-14MSD - 1	COMP-1-(0-1)MSD	22638	4.5149	-	7471B	PPB	20 Aug 2020 14:40:30	U	SMPL
L3729-15 - 1	COMP-2-(1-2)	3139	0.6198	-	7471B	PPB	20 Aug 2020 14:42:46	U	SMPL
L3729-16 - 1	COMP-3-(2-3)	1738	0.3399	-	7471B	PPB	20 Aug 2020 14:45:03	U	SMPL
L3729-17 - 1	COMP-4-(3-4)	5903	1.1719	-	7471B	PPB	20 Aug 2020 14:47:20	U	SMPL
L3729-18 - 1	COMP-5-(4-5)	2204	0.4330	-	7471B	PPB	20 Aug 2020 14:49:37	U	SMPL
L3742-03 - 1	TP-10-S	1502	0.2928	-	7471B	PPB	20 Aug 2020 14:51:52	U	SMPL
CCV57 - 1	CCV57	25998	5.1861	-	7471B	PPB	20 Aug 2020 14:54:08	U	SMPL
CCB57 - 1	CCB57	-103	-0.0278	-	7471B	PPB	20 Aug 2020 14:56:24	U	SMPL
L3742-07 - 1	TP-10-D	4373	0.8663	-	7471B	PPB	20 Aug 2020 14:58:40	U	SMPL
L3729-14LX5 - 1		633	0.1192	-	7471B	PPB	20 Aug 2020 15:00:55	U	SMPL
L3729-14A - 1		22426	4.4726	-	7471B	PPB	20 Aug 2020 15:07:23	U	SMPL
L3720-05DLX5 - 1		16939	3.3765	-	7471B	PPB	20 Aug 2020 15:09:38	U	SMPL
L3720-07DLX2 - 1		28941	5.7740	-	7471B	PPB	20 Aug 2020 15:11:54	U	SMPL
CCV58 - 1	CCV58	24498	4.8865	-	7471B	PPB	20 Aug 2020 15:14:11	U	SMPL
CCB58 - 1	CCB58	-105	-0.0282	-	7471B	PPB	20 Aug 2020 15:16:29	U	SMPL