



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB110626

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	08/20/2020	09:50
2	L3715-01		1	8	100.00	>212.0	>212.0	08/20/2020	10:20
3	L3715-01DUP		1	8	100.00	>212.0	>212.0	08/20/2020	10:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB110626

WorkList Name : fp82020 WorkList ID : 141642 Department : Wet-Chemistry Date : 08-20-2020 09:11:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/24/2020	Solid	L3715-01	A Flash Point	Cool 4 deg C	PSEG01	K42	226	08/17/2020	1010A

Date/Time 08/20/2020 09:20
 Received by: KS (wetchem)
 Relinquished by: KS (wetchem)

Date/Time 08/20/2020 10:50
 Received by: KS (wetchem)
 Relinquished by: KS (wetchem)