



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB110630

Slope : 97.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	08/20/2020	17:05
2	CAL2	1	Water	NA	NA	21.2	7.00	08/20/2020	17:06
3	CAL3	1	Water	NA	NA	21.2	10.03	08/20/2020	17:08
4	ICV	1	Water	NA	NA	21.1	6.98	08/20/2020	17:10
5	CCV1	1	Water	NA	NA	21.2	2.02	08/20/2020	17:11
6	L3506-10	1	Solid	20.03	20	23.6	8.59	08/20/2020	17:12
7	L3747-02	1	Solid	20.04	20	23.3	10.12	08/20/2020	17:15
8	L3747-02DUP	1	Solid	20.03	20	23.4	10.13	08/20/2020	17:16
9	CCV2	1	Water	NA	NA	21.1	12.02	08/20/2020	17:19

WORKLIST(Hardcopy Internal Chain)

VB 110630

WorkList Name : CORROSIONITY L3715 WorkList ID : 141654 Department : Wet-Chemistry Date : 08-20-2020 11:29:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/20/2020	Solid	L3506-10	Corrosivity	Cool 4 deg C	PSEG05		OR-03-082020	08/20/2020	9045D
08/20/2020	Solid	L3747-02	Corrosivity	Cool 4 deg C	PSEG01	K51	340	08/20/2020	9045D

Date/Time 08/20/2020 16:00 Chem
 Received by: Jg (wet)
 Relinquished by: DR (Digestion Lab)

Date/Time 08/20/2020 14:25
 Received by: DR (Digestion Lab)
 Relinquished by: Jg (wet) Chem