



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110631

Slope : 97.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	08/20/2020	17:05
2	CAL2	1	Water	NA	NA	21.2	7.00	08/20/2020	17:06
3	CAL3	1	Water	NA	NA	21.2	10.03	08/20/2020	17:08
4	ICV	1	Water	NA	NA	21.1	6.98	08/20/2020	17:10
5	CCV1	1	Water	NA	NA	21.2	2.02	08/20/2020	17:11
6	L3744-01	1	Water	NA	NA	24.3	4.09	08/20/2020	17:22
7	L3744-03	1	Water	NA	NA	24.2	4.15	08/20/2020	17:25
8	L3744-05	1	Water	NA	NA	24.0	4.64	08/20/2020	17:26
9	L3744-05DUP	1	Water	NA	NA	24.1	4.65	08/20/2020	17:27
10	CCV2	1	Water	NA	NA	21.2	12.02	08/20/2020	17:30

18 110631

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3747 WorkList ID : 141656 Department : Wet-Chemistry Date : 08-20-2020 16:10:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/20/2020	Water	L3744-01 ✓	pH	Cool 4 deg C	PSEG01	K61	40556	08/20/2020	9040C
08/20/2020	Water	L3744-03 ✓	pH	Cool 4 deg C	PSEG01	K61	40615	08/20/2020	9040C
08/20/2020	Water	L3744-05	pH	Cool 4 deg C	PSEG01	K61	16718	08/20/2020	9040C

Date/Time 08/20/2020 16:15
Received by: JR (Wet-Chem)
Relinquished by: DR (Digestion Lab)

Date/Time 08/20/2020 17:25
Received by: DR (Digestion Lab)
Relinquished by: JR (Wet-Chem)