

LB110641


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

35237



A= 0.0000e+000

B= 2.8620e-004

C= 3.0400e-002

Rho= 0.9997570

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.030	0.030	-2	0.000	-2				
0.2	0.200	0.210	0.010	628	0.0 %	628				
2.5	2.500	2.508	0.008	8657	0.0 %	8657				
5.0	5.000	4.996	-0.004	17350	0.0 %	17350				
7.5	7.500	7.341	-0.159	25543	0.0 %	25543				
10.0	10.000	10.115	0.115	35237	0.0 %	35237				

LB110641 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Aug 2020 12:01:45

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-2	-	0.0000	7470A	PPB	21 Aug 2020 12:12:05	S	Std
0.2 - 1	50.2	628	-	0.2000	7470A	PPB	21 Aug 2020 12:14:12	S	Std
2.5 - 1	52.5	8657	-	2.5000	7470A	PPB	21 Aug 2020 12:16:20	S	Std
5.0 - 1	55	17350	-	5.0000	7470A	PPB	21 Aug 2020 12:18:28	S	Std
7.5 - 1	57.5	25543	-	7.5000	7470A	PPB	21 Aug 2020 12:20:35	S	Std
10.0 - 1	59	35237	-	10.0000	7470A	PPB	21 Aug 2020 12:22:42	S	Std
ICV78 - 1	ICV78	14235	4.1045	-	7470A	PPB	21 Aug 2020 12:25:20	U	SMPL
ICB78 - 1	ICB78	-203	-0.0277	-	7470A	PPB	21 Aug 2020 12:27:25	U	SMPL
CCV62 - 1	CCV62	17720	5.1019	-	7470A	PPB	21 Aug 2020 12:29:31	U	SMPL
CCB62 - 1	CCB62	-230	-0.0354	-	7470A	PPB	21 Aug 2020 12:31:37	U	SMPL
CRA - 1	CRA	599	0.2018	-	7470A	PPB	21 Aug 2020 12:33:44	U	SMPL
HighStd - 1	HighStd	35056	10.0634	-	7470A	PPB	21 Aug 2020 12:35:50	U	SMPL
ChkStd - 1	ChkStd	25277	7.2647	-	7470A	PPB	21 Aug 2020 12:37:56	U	SMPL
PB131141BL - 1	PBW	-340	-0.0669	-	7470A	PPB	21 Aug 2020 12:40:05	U	SMPL
PB131141BS - 1	LCSW	13323	3.8434	-	7470A	PPB	21 Aug 2020 12:42:13	U	SMPL
L3506-10 - 1	OR-03-082020	-237	-0.0374	-	7470A	PPB	21 Aug 2020 12:44:20	U	SMPL
L3506-10DUP - 1	OR-03-082020DUP	-80	0.0075	-	7470A	PPB	21 Aug 2020 12:46:29	U	SMPL
L3506-10MS - 1	OR-03-082020MS	12864	3.7121	-	7470A	PPB	21 Aug 2020 12:48:36	U	SMPL
L3506-10MSD - 1	OR-03-082020MSD	13016	3.7556	-	7470A	PPB	21 Aug 2020 12:50:44	U	SMPL
L3715-01 - 1	226	-105	0.0003	-	7470A	PPB	21 Aug 2020 12:52:52	U	SMPL
CCV63 - 1	CCV63	17131	4.9333	-	7470A	PPB	21 Aug 2020 12:54:59	U	SMPL
CCB63 - 1	CCB63	-241	-0.0386	-	7470A	PPB	21 Aug 2020 12:57:05	U	SMPL
PB131141TB - 1	PB131120	-67	0.0112	-	7470A	PPB	21 Aug 2020 12:59:11	U	SMPL
L3506-10LX5 - 1		-42	0.0184	-	7470A	PPB	21 Aug 2020 13:01:17	U	SMPL
L3506-10A - 1		13770	3.9714	-	7470A	PPB	21 Aug 2020 13:03:23	U	SMPL
CCV64 - 1	CCV64	17315	4.9860	-	7470A	PPB	21 Aug 2020 13:05:29	U	SMPL
CCB64 - 1	CCB64	-253	-0.0420	-	7470A	PPB	21 Aug 2020 13:07:37	U	SMPL