



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB110645

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP86437
Turbidity Calibration std, 20NTU	WP86433
Turbidity Calibration std, 40NTU	WP86432
Turbidity Calibration std, 1NTU	WP86436
Turbidity Calibration std, 5NTU	WP86435
Turbidity Calibration - CCV std, 10 NTU	WP86434
10 NTU Standard 500 ml	W2489

Intercept: 0.2027

Slope: 0.9728

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.245	08/21/2020	11:15
2	CAL2	1	1	1.220	08/21/2020	11:17
3	CAL3	5	1	5.110	08/21/2020	11:20
4	CAL4	10	1	9.870	08/21/2020	11:22
5	CAL5	20	1	19.500	08/21/2020	11:25
6	CAL6	40	1	39.200	08/21/2020	11:27



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB110645

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.400	10.482	08/21/2020	11:30
2	ICB		1	0.262	0.061	08/21/2020	11:32
3	CCV1	10	1	9.930	9.999	08/21/2020	11:34
4	CCB1		1	0.246	0.045	08/21/2020	11:36
5	LB110645BL		1	0.268	0.067	08/21/2020	11:38
6	L3775-01		1	36.800	37.621	08/21/2020	11:40
7	L3775-01DUP		1	37.200	38.032	08/21/2020	11:42
8	CCV2	10	1	9.900	9.968	08/21/2020	11:45
9	CCB2		1	0.253	0.052	08/21/2020	11:47

LB110645
WORKLIST(Hardcopy Internal Chain)

WorkList Name : Turbidity-82120 WorkList ID : 141686 Department : Wet-Chemistry Date : 08-21-2020 10:47:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/21/2020	Water	L-3775-01	Turbidity	Cool 4 deg C	LIRO01	L12	UDS-4-26	08/19/2020	SM2130 B

Date/Time 8/21/2020 10:50
 Received by: AM(wet-chem)
 Relinquished by: JA (wet-chem)

Date/Time 8/21/2020 11:50
 Received by: JA (wet-chem)
 Relinquished by: AM(wet-chem)