


LB110647

7471B

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 1.9712e-004

C= 2.1448e-002

Rho= 0.9999295

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.031	0.031	47	0.000	47				
0.2	0.200	0.191	-0.009	859	0.0 %	859				
2.5	2.500	2.452	-0.048	12331	0.0 %	12331				
5.0	5.000	5.061	0.061	25568	0.0 %	25568				
7.5	7.500	7.439	-0.061	37629	0.0 %	37629				
10.0	10.000	10.026	0.026	50755	0.0 %	50755				

LB110647 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 21 Aug 2020 13:50:35

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	47	-	0.0000	7471B	PPB	21 Aug 2020 13:52:12	S	Std
0.2 - 1	50.2	859	-	0.2000	7471B	PPB	21 Aug 2020 13:54:29	S	Std
2.5 - 1	52.5	12331	-	2.5000	7471B	PPB	21 Aug 2020 13:56:46	S	Std
5.0 - 1	55	25568	-	5.0000	7471B	PPB	21 Aug 2020 13:59:03	S	Std
7.5 - 1	57.5	37629	-	7.5000	7471B	PPB	21 Aug 2020 14:01:20	S	Std
10.0 - 1	510	50755	-	10.0000	7471B	PPB	21 Aug 2020 14:06:15	S	Std
ICV79 - 1	ICV79	19403	3.8461	-	7471B	PPB	21 Aug 2020 14:08:57	U	SMPL
ICB79 - 1	ICB79	-157	-0.0095	-	7471B	PPB	21 Aug 2020 14:11:15	U	SMPL
CCV65 - 1	CCV65	25607	5.0691	-	7471B	PPB	21 Aug 2020 14:13:31	U	SMPL
CCB65 - 1	CCB65	-167	-0.0115	-	7471B	PPB	21 Aug 2020 14:15:46	U	SMPL
CRA - 1	CRA	831	0.1853	-	7471B	PPB	21 Aug 2020 14:18:03	U	SMPL
HighStd - 1	HighStd	50277	9.9320	-	7471B	PPB	21 Aug 2020 14:20:18	U	SMPL
ChkStd - 1	ChkStd	38048	7.5214	-	7471B	PPB	21 Aug 2020 14:22:33	U	SMPL
PB131126BL - 1	PBS	-300	-0.0377	-	7471B	PPB	21 Aug 2020 14:24:52	U	SMPL
PB131126BS - 1	LCSS	17555	3.4819	-	7471B	PPB	21 Aug 2020 14:27:10	U	SMPL
L3506-06 - 1	OR-03-082020	11915	2.3701	-	7471B	PPB	21 Aug 2020 14:29:26	U	SMPL
L3734-02 - 1	EROC-BF-1A	4146	0.8387	-	7471B	PPB	21 Aug 2020 14:31:44	U	SMPL
L3734-03 - 1	EROC-BF-1B	4092	0.8281	-	7471B	PPB	21 Aug 2020 14:34:02	U	SMPL
L3734-04 - 1	EROC-BF-1C	5488	1.1032	-	7471B	PPB	21 Aug 2020 14:36:19	U	SMPL
L3734-05 - 1	EROC-BF-1D	4627	0.9335	-	7471B	PPB	21 Aug 2020 14:38:35	U	SMPL
CCV66 - 1	CCV66	24915	4.9327	-	7471B	PPB	21 Aug 2020 14:40:51	U	SMPL
CCB66 - 1	CCB66	-187	-0.0154	-	7471B	PPB	21 Aug 2020 14:43:06	U	SMPL
L3734-06 - 1	EROC-BF-1E	7962	1.5909	-	7471B	PPB	21 Aug 2020 14:45:23	U	SMPL
L3734-07 - 1	EROC-BF-1F	6379	1.2789	-	7471B	PPB	21 Aug 2020 14:47:38	U	SMPL
L3734-07DUP - 1	EROC-BF-1FDUP	5915	1.1874	-	7471B	PPB	21 Aug 2020 14:49:53	U	SMPL
L3734-07MS - 1	EROC-BF-1FMS	25926	5.1319	-	7471B	PPB	21 Aug 2020 14:58:32	U	SMPL
L3734-07MSD - 1	EROC-BF-1FMDS	24910	4.9317	-	7471B	PPB	21 Aug 2020 15:00:48	U	SMPL
L3746-02 - 1	359	5441	1.0940	-	7471B	PPB	21 Aug 2020 15:03:06	U	SMPL
L3746-04 - 1	224	3858	0.7819	-	7471B	PPB	21 Aug 2020 15:05:24	U	SMPL
L3746-06 - 1	212	2914	0.5959	-	7471B	PPB	21 Aug 2020 15:07:42	U	SMPL
L3747-01 - 1	340	1236	0.2651	-	7471B	PPB	21 Aug 2020 15:09:59	U	SMPL
CCV67 - 1	CCV67	25081	4.9654	-	7471B	PPB	21 Aug 2020 15:12:15	U	SMPL
CCB67 - 1	CCB67	-119	-0.0020	-	7471B	PPB	21 Aug 2020 15:14:30	U	SMPL
L3734-07LX5 - 1		1217	0.2613	-	7471B	PPB	21 Aug 2020 15:16:47	U	SMPL
L3734-07A - 1		24961	4.9417	-	7471B	PPB	21 Aug 2020 15:19:02	U	SMPL
CCV68 - 1	CCV68	25158	4.9806	-	7471B	PPB	21 Aug 2020 15:24:36	U	SMPL
CCB68 - 1	CCB68	-164	-0.0109	-	7471B	PPB	21 Aug 2020 15:26:51	U	SMPL