



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110657

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	3.99	08/21/2020	17:20
2	CAL2	1	Water	NA	NA	21.2	7.00	08/21/2020	17:21
3	CAL3	1	Water	NA	NA	21.1	10.03	08/21/2020	17:22
4	ICV	1	Water	NA	NA	21.2	7.00	08/21/2020	17:25
5	CCV1	1	Water	NA	NA	21.1	2.01	08/21/2020	17:26
6	L3771-01	1	Water	NA	NA	26.4	7.41	08/21/2020	17:28
7	L3771-01DUP	1	Water	NA	NA	26.5	7.42	08/21/2020	17:29
8	L3774-06	1	Water	NA	NA	25.2	8.44	08/21/2020	17:30
9	CCV2	1	Water	NA	NA	21.2	12.02	08/21/2020	17:32

JB 110657

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH WL3774 WorkList ID : 141693 Department : Wet-Chemistry Date : 08-21-2020 15:14:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/21/2020	Water	L3771-01	pH	Cool 4 deg C	PSEG01	L11	DRUM-1	08/21/2020	9040C
08/21/2020	Water	L3774-06	pH	Cool 4 deg C	PSEG01	L12	TOTE-1-2	08/21/2020	9040C

Date/Time 08.21.2020 16:00
 Received by: JO [Signature]
 Relinquished by: JO [Signature]

Date/Time 08.21.2020 18:00
 Received by: JO [Signature]
 Relinquished by: JO [Signature]