



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110659

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	3.99	08/21/2020	17:20
2	CAL2	1	Water	NA	NA	21.2	7.00	08/21/2020	17:21
3	CAL3	1	Water	NA	NA	21.1	10.03	08/21/2020	17:22
4	ICV	1	Water	NA	NA	21.2	7.00	08/21/2020	17:25
5	CCV1	1	Water	NA	NA	21.1	2.01	08/21/2020	17:26
6	L3774-04	1	Solid	20.03	20	24.0	8.79	08/21/2020	17:50
7	L3774-04DUP	1	Solid	20.04	20	24.1	8.80	08/21/2020	17:51
8	CCV2	1	Water	NA	NA	21.1	12.02	08/21/2020	18:00

VR110659

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S L3774 WorkList ID : 141692 Department : Wet-Chemistry Date : 08-21-2020 15:14:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/21/2020	Solid	L3774-04	pH	Cool 4 deg C	PSEG01	L12	DRUM-2	08/21/2020	9045D

Date/Time 08/21/2020 16:00
Received by: [Signature]
Relinquished by: [Signature]

Date/Time 08/21/2020 18:00
Received by: [Signature]
Relinquished by: [Signature]