


LB110667

7471B

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.014	0.014	20	0.000	20				
0.2	0.200	0.195	-0.005	908	0.0 %	908				
2.5	2.500	2.539	0.039	12404	0.0 %	12404				
5.0	5.000	4.855	-0.145	23763	0.0 %	23763				
7.5	7.500	7.663	0.163	37538	0.0 %	37538				
10.0	10.000	9.935	-0.065	48680	0.0 %	48680				

LB110667 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 24 Aug 2020 14:21:03				
Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	SO	20	-	0.00007471B	PPB	24 Aug 2020 14:28:09	S	Std
0.2 - 1	SO.2	908	-	0.20007471B	PPB	24 Aug 2020 14:30:25	S	Std
2.5 - 1	SO.5	12404	-	2.50007471B	PPB	24 Aug 2020 14:32:42	S	Std
5.0 - 1	SO	23763	-	5.00007471B	PPB	24 Aug 2020 14:35:00	S	Std
7.5 - 1	SO.5	37538	-	7.50007471B	PPB	24 Aug 2020 14:37:17	S	Std
10.0 - 1	SO	48680	-	10.00007471B	PPB	24 Aug 2020 14:42:08	S	Std
ICV81 - 1	ICV81	19840	4.0548	-7471B	PPB	24 Aug 2020 14:44:50	U	SMPL
ICB81 - 1	ICB81	-152	-0.0215	-7471B	PPB	24 Aug 2020 14:47:06	U	SMPL
CCV72 - 1	CCV72	24139	4.9313	-7471B	PPB	24 Aug 2020 14:49:23	U	SMPL
CCB72 - 1	CCB72	-176	-0.0264	-7471B	PPB	24 Aug 2020 14:51:37	U	SMPL
CRA - 1	CRA	896	0.1921	-7471B	PPB	24 Aug 2020 14:53:54	U	SMPL
HighStd - 1	HighStd	48547	9.9081	-7471B	PPB	24 Aug 2020 14:56:10	U	SMPL
ChkStd - 1	ChkStd	36423	7.4360	-7471B	PPB	24 Aug 2020 14:58:26	U	SMPL
PB131161BL - 1	PBS	-295	-0.0507	-7471B	PPB	24 Aug 2020 15:00:44	U	SMPL
PB131161BS - 1	LCSS	17539	3.5856	-7471B	PPB	24 Aug 2020 15:03:02	U	SMPL
L3772-02 - 1	349-346	24692	5.0441	-7471B	PPB	24 Aug 2020 15:05:18	U	SMPL
L3772-05 - 1	202-986	54559	11.1339	-7471B	PPB	24 Aug 2020 15:07:35	U	SMPL
L3773-01 - 1	DRUM-COMP	483	0.1079	-7471B	PPB	24 Aug 2020 15:09:55	U	SMPL
L3773-03 - 1	OILY-DEBRIS	105	0.0309	-7471B	PPB	24 Aug 2020 15:12:14	U	SMPL
L3780-01 - 1	JC-03-082120-A	9033	1.8513	-7471B	PPB	24 Aug 2020 15:14:30	U	SMPL
CCV73 - 1	CCV73	24514	5.0078	-7471B	PPB	24 Aug 2020 15:16:45	U	SMPL
CCB73 - 1	CCB73	-156	-0.0224	-7471B	PPB	24 Aug 2020 15:19:02	U	SMPL
L3780-09 - 1	JC-03-082120-B	7713	1.5821	-7471B	PPB	24 Aug 2020 15:21:18	U	SMPL
L3780-17 - 1	JC-03-082120-C	11229	2.2990	-7471B	PPB	24 Aug 2020 15:23:34	U	SMPL
L3780-17DUP - 1	JC-03-082120-CDUP	11733	2.4018	-7471B	PPB	24 Aug 2020 15:25:50	U	SMPL
L3780-17MS - 1	JC-03-082120-CMS	32364	6.6084	-7471B	PPB	24 Aug 2020 15:28:06	U	SMPL
L3780-17MSD - 1	JC-03-082120-CMSD	31875	6.5087	-7471B	PPB	24 Aug 2020 15:30:21	U	SMPL
L3780-17LX5 - 1		1828	0.3822	-7471B	PPB	24 Aug 2020 15:32:39	U	SMPL
L3780-17A - 1		30131	6.1531	-7471B	PPB	24 Aug 2020 15:34:58	U	SMPL
L3772-05DLX2 - 1		26866	5.4874	-7471B	PPB	24 Aug 2020 15:37:15	U	SMPL
CCV74 - 1	CCV74	24195	4.9428	-7471B	PPB	24 Aug 2020 15:42:06	U	SMPL
CCB74 - 1	CCB74	-117	-0.0144	-7471B	PPB	24 Aug 2020 15:44:21	U	SMPL