



## Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB110668

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

| Reagent/Standard | Lot/Log # |
|------------------|-----------|
| p-xylene (ICV)   | W2168     |

| Seq | LabID       | True Value °F | DL | Initial Sample °C | Celsius °C | Result °F | Final Result °F | Anal Date  | Anal Time |
|-----|-------------|---------------|----|-------------------|------------|-----------|-----------------|------------|-----------|
| 1   | ICV         | 81            | 1  | 7                 | 27.00      | 80.6      | 80.6            | 08/24/2020 | 09:25     |
| 2   | L3771-01    |               | 1  | 8                 | 100.00     | >212.0    | >212.0          | 08/24/2020 | 09:50     |
| 3   | L3771-01DUP |               | 1  | 8                 | 100.00     | >212.0    | >212.0          | 08/24/2020 | 10:20     |

Result = (Celsius \* 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) \* 0.06

# WORKLIST(Hardcopy Internal Chain)

LB110668

WorkList Name : fp82420      WorkList ID : 141712      Department : Wet-Chemistry      Date : 08-24-2020 08:58:01

| Due Date   | Matrix | Sample   | Test        | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|----------|-------------|--------------|----------|------------------|-----------------|--------------|--------|
| 08/28/2020 | Water  | L3771-01 | Flash Point | Cool 4 deg C | PSEG01   | L11              | DRUM-1          | 08/21/2020   | 1010A  |

Date/Time 08/24/2020 09:10  
 Received by: KS (Weber)  
 Relinquished by: KS (Weber)

Date/Time 08/24/2020 10:25  
 Received by: KS (Weber)  
 Relinquished by: KS (Weber)