



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110669

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	08/24/2020	16:10
2	CAL2	1	Water	NA	NA	21.2	7.01	08/24/2020	16:11
3	CAL3	1	Water	NA	NA	21.2	10.03	08/24/2020	16:12
4	ICV	1	Water	NA	NA	21.2	6.98	08/24/2020	16:15
5	CCV1	1	Water	NA	NA	21.2	2.02	08/24/2020	16:16
6	L3787-02	1	Water	NA	NA	25.6	6.64	08/24/2020	16:19
7	L3787-03	1	Water	NA	NA	26.3	6.63	08/24/2020	16:20
8	L3788-03	1	Water	NA	NA	27.8	5.34	08/24/2020	16:22
9	L3788-03DUP	1	Water	NA	NA	27.9	5.35	08/24/2020	16:23
10	CCV2	1	Water	NA	NA	21.3	12.02	08/24/2020	16:25

LB 110669

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3788 W

WorkList ID : 141729

Department : Wet-Chemistry

Date : 08-24-2020 14:40:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/21/2020	Solid	L3774-09	pH	Cool 4 deg C	PSEG01	L12	TOTE-1-2	08/21/2020	9045D
08/24/2020	Water	L3787-02	pH	Cool 4 deg C	PSEG01	L13	1574	08/24/2020	9040C
08/24/2020	Water	L3787-03	pH	Cool 4 deg C	PSEG01	L13	1498	08/24/2020	9040C
08/24/2020	Water	L3788-03	pH	Cool 4 deg C	PSEG01	L13	1794	08/24/2020	9040C

Date/Time 08/24/2020 16:00

Received by: JP (Wet Chem)

Relinquished by: KS (Wet Chem)

Date/Time 08/24/2020

Received by: KS (Wet Chem)

Relinquished by: JP (Wet Chem)