


LB110681

7470A

INSTRUMENT ID : CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.008	-0.008	2	0.000	2				
0.2	0.200	0.217	0.017	757	0.0 %	757				
2.5	2.500	2.478	-0.022	8324	0.0 %	8324				
5.0	5.000	5.120	0.120	17170	0.0 %	17170				
7.5	7.500	7.314	-0.186	24516	0.0 %	24516				
10.0	10.000	10.079	0.079	33770	0.0 %	33770				

LB110681 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 25 Aug 2020 10:22:36

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	2	-	0.00007470A	PPB		25 Aug 2020 10:28:39	S	Std
0.2 - 1	50.2	757	-	0.20007470A	PPB		25 Aug 2020 10:30:47	S	Std
2.5 - 1	52.5	8324	-	2.50007470A	PPB		25 Aug 2020 10:32:54	S	Std
5.0 - 1	55	17170	-	5.00007470A	PPB		25 Aug 2020 10:35:02	S	Std
7.5 - 1	57.5	24516	-	7.50007470A	PPB		25 Aug 2020 10:39:36	S	Std
10.0 - 1	510	33770	-	10.00007470A	PPB		25 Aug 2020 10:46:35	S	Std
ICV82 - 1	ICV82	12821	3.8210	-7470A	PPB		25 Aug 2020 10:49:08	U	SMPL
ICB82 - 1	ICB82	-36	-0.0196	-7470A	PPB		25 Aug 2020 10:53:45	U	SMPL
CCV75 - 1	CCV75	16129	4.8091	-7470A	PPB		25 Aug 2020 10:55:51	U	SMPL
CCB75 - 1	CCB75	-34	-0.0190	-7470A	PPB		25 Aug 2020 11:00:22	U	SMPL
CRA - 1	CRA	683	0.1952	-7470A	PPB		25 Aug 2020 11:02:28	U	SMPL
HighStd - 1	HighStd	33528	10.0065	-7470A	PPB		25 Aug 2020 11:04:33	U	SMPL
ChkStd - 1	ChkStd	22594	6.7403	-7470A	PPB		25 Aug 2020 11:06:40	U	SMPL
PB131169BL - 1	PBW	60	0.0091	-7470A	PPB		25 Aug 2020 11:11:14	U	SMPL
PB131169BS - 1	LCSW	11781	3.5103	-7470A	PPB		25 Aug 2020 11:13:20	U	SMPL
L3774-01 - 1	EXTERIOR-BUILDING	-87	-0.0348	-7470A	PPB		25 Aug 2020 11:15:27	U	SMPL
L3774-02 - 1	DRUM-1	-13	-0.0127	-7470A	PPB		25 Aug 2020 11:17:34	U	SMPL
L3774-03 - 1	TOTE-1-1	383	0.1056	-7470A	PPB		25 Aug 2020 11:19:42	U	SMPL
L3774-04 - 1	DRUM-2	72	0.0127	-7470A	PPB		25 Aug 2020 11:21:49	U	SMPL
L3774-09 - 1	TOTE-1-2	103	0.0219	-7470A	PPB		25 Aug 2020 11:23:55	U	SMPL
CCV76 - 1	CCV76	17114	5.1034	-7470A	PPB		25 Aug 2020 11:26:02	U	SMPL
CCB76 - 1	CCB76	-191	-0.0659	-7470A	PPB		25 Aug 2020 11:28:08	U	SMPL
L3780-08 - 1	JC-03-082120-A	8	-0.0065	-7470A	PPB		25 Aug 2020 11:30:14	U	SMPL
L3780-16 - 1	JC-03-082120-B	41	0.0034	-7470A	PPB		25 Aug 2020 11:32:21	U	SMPL
L3780-23 - 1	JC-03-082120-C	121	0.0273	-7470A	PPB		25 Aug 2020 11:34:28	U	SMPL
L3780-23DUP - 1	JC-03-082120-CDUP	169	0.0416	-7470A	PPB		25 Aug 2020 11:36:34	U	SMPL
L3780-23MS - 1	JC-03-082120-CMS	11897	3.5450	-7470A	PPB		25 Aug 2020 11:38:41	U	SMPL
L3780-23MSD - 1	JC-03-082120-CMSD	10993	3.2749	-7470A	PPB		25 Aug 2020 11:40:48	U	SMPL
PB131170BL - 1	PBW	46	0.0049	-7470A	PPB		25 Aug 2020 11:42:55	U	SMPL
PB131170BS - 1	LCSW	11831	3.5253	-7470A	PPB		25 Aug 2020 11:49:41	U	SMPL
L3771-02 - 1	DRUM-1	-147	-0.0528	-7470A	PPB		25 Aug 2020 11:51:48	U	SMPL
L3771-02DUP - 1	DRUM-1DUP	-8	-0.0112	-7470A	PPB		25 Aug 2020 11:53:55	U	SMPL
CCV77 - 1	CCV77	17449	5.2034	-7470A	PPB		25 Aug 2020 11:56:01	U	SMPL
CCB77 - 1	CCB77	-215	-0.0731	-7470A	PPB		25 Aug 2020 11:58:07	U	SMPL
L3771-02MS - 1	DRUM-1MS	10854	3.2334	-7470A	PPB		25 Aug 2020 12:00:14	U	SMPL
L3771-02MSD - 1	DRUM-1MSD	11031	3.2863	-7470A	PPB		25 Aug 2020 12:04:55	U	SMPL
PB131169TB - 1	PB131150TB	-158	-0.0560	-7470A	PPB		25 Aug 2020 12:07:00	U	SMPL
PB131170TB - 1	PB131166TB	9	-0.0062	-7470A	PPB		25 Aug 2020 12:09:09	U	SMPL
L3780-23LX5 - 1		99	0.0207	-7470A	PPB		25 Aug 2020 12:11:16	U	SMPL
L3780-23A - 1		12645	3.7684	-7470A	PPB		25 Aug 2020 12:13:23	U	SMPL
L3771-02LX5 - 1		-66	-0.0286	-7470A	PPB		25 Aug 2020 12:15:30	U	SMPL
L3771-02A - 1		12493	3.7230	-7470A	PPB		25 Aug 2020 12:17:38	U	SMPL
CCV78 - 1	CCV78	16458	4.9074	-7470A	PPB		25 Aug 2020 12:19:46	U	SMPL
CCB78 - 1	CCB78	-139	-0.0504	-7470A	PPB		25 Aug 2020 12:21:53	U	SMPL