


LB110684

7471B

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.057	0.057	93	0.000	93				
0.2	0.200	0.203	0.003	826	0.0 %	826				
2.5	2.500	2.464	-0.036	12212	0.0 %	12212				
5.0	5.000	4.925	-0.075	24602	0.0 %	24602				
7.5	7.500	7.522	0.022	37678	0.0 %	37678				
10.0	10.000	10.028	0.028	50294	0.0 %	50294				

LB110684 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 25 Aug 2020 13:26:37

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	93	-	0.0000	7471B	PPB	25 Aug 2020 13:37:42	S	Std
0.2 - 1	50.2	826	-	0.2000	7471B	PPB	25 Aug 2020 13:39:59	S	Std
2.5 - 1	52.5	12212	-	2.5000	7471B	PPB	25 Aug 2020 13:42:16	S	Std
5.0 - 1	55	24602	-	5.0000	7471B	PPB	25 Aug 2020 13:44:33	S	Std
7.5 - 1	57.5	37678	-	7.5000	7471B	PPB	25 Aug 2020 13:46:50	S	Std
10.0 - 1	50	50294	-	10.0000	7471B	PPB	25 Aug 2020 13:49:07	S	Std
ICV83 - 1	ICV83	19328	3.8777	-	7471B	PPB	25 Aug 2020 13:52:36	U	SMPL
ICB83 - 1	ICB83	-80	0.0227	-	7471B	PPB	25 Aug 2020 13:54:51	U	SMPL
CCV79 - 1	CCV79	25003	5.0049	-	7471B	PPB	25 Aug 2020 13:57:07	U	SMPL
CCB79 - 1	CCB79	-103	0.0182	-	7471B	PPB	25 Aug 2020 13:59:22	U	SMPL
CRA - 1	CRA	812	0.1999	-	7471B	PPB	25 Aug 2020 14:01:38	U	SMPL
HighStd - 1	HighStd	50273	10.0241	-	7471B	PPB	25 Aug 2020 14:03:54	U	SMPL
ChkStd - 1	ChkStd	33995	6.7909	-	7471B	PPB	25 Aug 2020 14:06:09	U	SMPL
PB131191BL - 1	PBS	-229	-0.0069	-	7471B	PPB	25 Aug 2020 14:08:27	U	SMPL
PB131191BS - 1	LCSS	17429	3.5005	-	7471B	PPB	25 Aug 2020 14:10:45	U	SMPL
L3784-01 - 1	RO-10	10908	2.2052	-	7471B	PPB	25 Aug 2020 14:13:01	U	SMPL
L3784-01DUP - 1	RO-10DUP	10709	2.1657	-	7471B	PPB	25 Aug 2020 14:15:19	U	SMPL
L3784-01MS - 1	RO-10MS	29871	5.9718	-	7471B	PPB	25 Aug 2020 14:17:38	U	SMPL
L3784-01MSD - 1	RO-10MSD	30976	6.1913	-	7471B	PPB	25 Aug 2020 14:19:56	U	SMPL
L3784-04 - 1	RO-11	2797	0.5942	-	7471B	PPB	25 Aug 2020 14:22:13	U	SMPL
CCV80 - 1	CCV80	24673	4.9393	-	7471B	PPB	25 Aug 2020 14:24:30	U	SMPL
CCB80 - 1	CCB80	-130	0.0128	-	7471B	PPB	25 Aug 2020 14:26:45	U	SMPL
L3784-07 - 1	RO-23	2261	0.4877	-	7471B	PPB	25 Aug 2020 14:29:02	U	SMPL
L3784-10 - 1	RO-24	2822	0.5992	-	7471B	PPB	25 Aug 2020 14:31:17	U	SMPL
L3784-13 - 1	RO-06	6966	1.4223	-	7471B	PPB	25 Aug 2020 14:33:33	U	SMPL
L3784-16 - 1	RO-04	1186	0.2742	-	7471B	PPB	25 Aug 2020 14:35:49	U	SMPL
L3785-01 - 1	CHRT-09	1575	0.3515	-	7471B	PPB	25 Aug 2020 14:38:05	U	SMPL
L3785-04 - 1	CHRT-18	1467	0.3300	-	7471B	PPB	25 Aug 2020 14:40:21	U	SMPL
L3784-01LX5 - 1		1652	0.3668	-	7471B	PPB	25 Aug 2020 14:42:37	U	SMPL
L3784-01A - 1		28273	5.6544	-	7471B	PPB	25 Aug 2020 14:44:54	U	SMPL
CCV81 - 1	CCV81	25053	5.0148	-	7471B	PPB	25 Aug 2020 14:47:11	U	SMPL
CCB81 - 1	CCB81	-97	0.0194	-	7471B	PPB	25 Aug 2020 14:49:28	U	SMPL