



### Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110685

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	3.99	08/25/2020	13:05
2	CAL2	1	Water	NA	NA	21.3	7.00	08/25/2020	13:06
3	CAL3	1	Water	NA	NA	21.3	10.03	08/25/2020	13:07
4	ICV	1	Water	NA	NA	21.3	7.01	08/25/2020	13:09
5	CCV1	1	Water	NA	NA	21.3	2.02	08/25/2020	13:11
6	L3792-01	1	Water	NA	NA	26.5	4.24	08/25/2020	13:12
7	L3792-01DUP	1	Water	NA	NA	26.6	4.25	08/25/2020	13:13
8	L3792-02	1	Water	NA	NA	26.5	8.35	08/25/2020	13:15
9	L3792-06	1	Water	NA	NA	26.5	7.70	08/25/2020	13:19
10	L3792-08	1	Water	NA	NA	26.7	7.91	08/25/2020	13:20
11	L3792-10	1	Water	NA	NA	26.8	8.50	08/25/2020	13:24
12	CCV2	1	Water	NA	NA	21.2	12.01	08/25/2020	13:25

LB-110685

### WORKLIST(Hardcopy Internal Chain)

WorkList Name: PH L3792 W      WorkList ID: 141762      Department: Wet-Chemistry      Date: 08-25-2020 12:59:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/25/2020	Water	L3792-01 <u>E</u>	pH	Cool 4 deg C	PSEG01	L21	40617	08/25/2020	9040C
08/25/2020	Water	L3792-02 <u>E</u>	pH	Cool 4 deg C	PSEG01	L21	40616	08/25/2020	9040C
08/25/2020	Water	L3792-06 <u>E</u>	pH	Cool 4 deg C	PSEG01	L21	COMP-1	08/25/2020	9040C
08/25/2020	Water	L3792-08 <u>E</u>	pH	Cool 4 deg C	PSEG01	L21	COMP-2	08/25/2020	9040C
08/25/2020	Water	L3792-10 <u>E</u>	pH	Cool 4 deg C	PSEG01	L21	COMP-3	08/25/2020	9040C

Date/Time 08/25/2020 13:00  
 Received by: J.P. Wilson  
 Relinquished by: J.P. Wilson

Date/Time 08/25/2020 15:00  
 Received by: J.P. Wilson  
 Relinquished by: J.P. Wilson