



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110686

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	3.99	08/25/2020	13:05
2	CAL2	1	Water	NA	NA	21.3	7.00	08/25/2020	13:06
3	CAL3	1	Water	NA	NA	21.3	10.03	08/25/2020	13:07
4	ICV	1	Water	NA	NA	21.3	7.01	08/25/2020	13:09
5	CCV1	1	Water	NA	NA	21.3	2.02	08/25/2020	13:11
6	L3792-13	1	Solid	20.03	20	24.7	6.53	08/25/2020	13:30
7	L3792-13DUP	1	Solid	20.04	20	24.8	6.54	08/25/2020	13:31
8	CCV2	1	Water	NA	NA	21.3	12.02	08/25/2020	13:33

1B110686

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S L3792 WorkList ID : 141763 Department : Wet-Chemistry Date : 08-25-2020 13:00:19

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/25/2020	Solid	L3792-13-B	pH	Cool 4 deg C	PSEG01	L21	40958	08/25/2020	9045D

Date/Time 08/25/2020 13:00
Received by: JP (wet chem)
Relinquished by: JP (wet)

Date/Time 08/25/2020 15:00
Received by: JP (wet)
Relinquished by: JP (wet) chem