



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Alexander

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB110698

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH12 (500ml)	W2616
Buffer Solution, PH2 (500ml)	W2592

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.7	4.00	08/26/2020	11:02
2	CAL2	1	Water	NA	NA	22.7	7.01	08/26/2020	11:04
3	CAL3	1	Water	NA	NA	22.7	10.03	08/26/2020	11:06
4	ICV	1	Water	NA	NA	22.7	6.98	08/26/2020	11:08
5	CCV1	1	Water	NA	NA	22.8	2.01	08/26/2020	11:10
6	L3805-02	1	Water	NA	NA	21.8	7.90	08/26/2020	11:13
7	L3805-02DUP	1	Water	NA	NA	21.8	7.92	08/26/2020	11:15
8	CCV2	1	Water	NA	NA	22.8	12.02	08/26/2020	11:17

LB110698
WORKLIST(Hardcopy Internal Chain)

WorkList Name : 82620-corrosivity-water WorkList ID : 141780 Department : Wet-Chemistry Date : 08-26-2020 09:47:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/26/2020	Water	L3805-02 B	Corrosivity	Cool 4 deg C	PSEG01	L31	FM782	08/26/2020	9040C

Date/Time 8/26/2020 10:50
Received by: AM (wet-chem)
Relinquished by: JF (SW)

Date/Time 8/26/2020 11:30
Received by: JF (SW)
Relinquished by: AM (wet-chem)