


LB110702

7471B

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.072	0.072	427	0.000	427				
0.2	0.200	0.214	0.014	1128	0.0 %	1128				
2.5	2.500	2.463	-0.037	12235	0.0 %	12235				
5.0	5.000	4.900	-0.100	24269	0.0 %	24269				
7.5	7.500	7.476	-0.024	36989	0.0 %	36989				
10.0	10.000	10.074	0.074	49819	0.0 %	49819				

LB110702 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Aug 2020 15:09:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	427	-	0.0000	7471B	PPB	26 Aug 2020 15:12:37	S	Std
0.2 - 1	57.2	1128	-	0.2000	7471B	PPB	26 Aug 2020 15:14:54	S	Std
2.5 - 1	52.5	12235	-	2.5000	7471B	PPB	26 Aug 2020 15:17:11	S	Std
5.0 - 1	55	24269	-	5.0000	7471B	PPB	26 Aug 2020 15:19:28	S	Std
7.5 - 1	57.5	36989	-	7.5000	7471B	PPB	26 Aug 2020 15:21:45	S	Std
10.0 - 1	57.5	49819	-	10.0000	7471B	PPB	26 Aug 2020 15:24:02	S	Std
ICV85 - 1	ICV85	18929	3.8188	-	7471B	PPB	26 Aug 2020 15:26:56	U	SMPL
ICB85 - 1	ICB85	284	0.0430	-	7471B	PPB	26 Aug 2020 15:29:11	U	SMPL
CCV86 - 1	CCV86	24829	5.0137	-	7471B	PPB	26 Aug 2020 15:31:27	U	SMPL
CCB86 - 1	CCB86	132	0.0123	-	7471B	PPB	26 Aug 2020 15:33:42	U	SMPL
CRA - 1	CRA	1128	0.2140	-	7471B	PPB	26 Aug 2020 15:35:59	U	SMPL
HighStd - 1	HighStd	49963	10.1035	-	7471B	PPB	26 Aug 2020 15:38:14	U	SMPL
ChkStd - 1	ChkStd	34645	7.0015	-	7471B	PPB	26 Aug 2020 15:40:29	U	SMPL
PB131213BL - 1	PBS	-26	-0.0197	-	7471B	PPB	26 Aug 2020 15:42:47	U	SMPL
PB131213BS - 1	LCSS	18097	3.6504	-	7471B	PPB	26 Aug 2020 15:47:37	U	SMPL
L3793-02 - 1	TP-1	902	0.1682	-	7471B	PPB	26 Aug 2020 15:54:44	U	SMPL
L3793-02DUP - 1	TP-1DUP	948	0.1775	-	7471B	PPB	26 Aug 2020 15:59:37	U	SMPL
L3793-02MS - 1	TP-1MS	18223	3.6759	-	7471B	PPB	26 Aug 2020 16:01:54	U	SMPL
L3793-02MSD - 1	TP-1MSD	18949	3.8229	-	7471B	PPB	26 Aug 2020 16:04:12	U	SMPL
L3793-05 - 1	TP-2	370	0.0605	-	7471B	PPB	26 Aug 2020 16:06:28	U	SMPL
CCV87 - 1	CCV87	25365	5.1222	-	7471B	PPB	26 Aug 2020 16:08:44	U	SMPL
CCB87 - 1	CCB87	21	-0.0102	-	7471B	PPB	26 Aug 2020 16:10:59	U	SMPL
L3793-08 - 1	TP-3	653	0.1178	-	7471B	PPB	26 Aug 2020 16:13:15	U	SMPL
L3793-11 - 1	TP-4	738	0.1350	-	7471B	PPB	26 Aug 2020 16:15:31	U	SMPL
L3793-14 - 1	TP-5	740	0.1354	-	7471B	PPB	26 Aug 2020 16:17:46	U	SMPL
L3793-17 - 1	TP-6	408	0.0682	-	7471B	PPB	26 Aug 2020 16:20:02	U	SMPL
L3793-20 - 1	TP-7	363	0.0590	-	7471B	PPB	26 Aug 2020 16:22:18	U	SMPL
L3793-23 - 1	TP-8	495	0.0858	-	7471B	PPB	26 Aug 2020 16:24:34	U	SMPL
L3793-26 - 1	TP-9	618	0.1107	-	7471B	PPB	26 Aug 2020 16:26:50	U	SMPL
L3794-01 - 1	336	2084	0.4076	-	7471B	PPB	26 Aug 2020 16:29:08	U	SMPL
L3793-02LX5 - 1		470	0.0807	-	7471B	PPB	26 Aug 2020 16:31:24	U	SMPL
L3793-02A - 1		20884	4.2147	-	7471B	PPB	26 Aug 2020 16:33:40	U	SMPL
CCV88 - 1	CCV88	25476	5.1447	-	7471B	PPB	26 Aug 2020 16:35:56	U	SMPL
CCB88 - 1	CCB88	77	0.0011	-	7471B	PPB	26 Aug 2020 16:38:11	U	SMPL