



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB110703

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	08/26/2020	12:35
2	L3792-01		1	8	100.00	>212.0	>212.0	08/26/2020	13:05
3	L3792-01DUP		1	8	100.00	>212.0	>212.0	08/26/2020	13:40
4	L3792-02		1	9	100.00	>212.0	>212.0	08/26/2020	14:10
5	L3792-06		1	8	100.00	>212.0	>212.0	08/26/2020	14:35
6	L3792-08		1	9	100.00	>212.0	>212.0	08/26/2020	15:05
7	L3792-10		1	8	100.00	>212.0	>212.0	08/26/2020	15:30
8	L3805-02		1	9	100.00	>212.0	>212.0	08/26/2020	15:55

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 110703

WorkList Name : fp82620 WorkList ID : 141787 Department : Wet-Chemistry Date : 08-26-2020 11:56:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/28/2020	Water	L3792-01	F	Cool 4 deg C	PSEG01	L21	40617	08/25/2020	1010A
08/28/2020	Water	L3792-02	B	Cool 4 deg C	PSEG01	L21	40616	08/25/2020	1010A
08/28/2020	Water	L3792-06	F	Cool 4 deg C	PSEG01	L21	COMP-1	08/25/2020	1010A
08/28/2020	Water	L3792-08	F	Cool 4 deg C	PSEG01	L21	COMP-2	08/25/2020	1010A
08/28/2020	Water	L3792-10	F	Cool 4 deg C	PSEG01	L21	COMP-3	08/25/2020	1010A
09/02/2020	Water	L3805-02	C	Cool 4 deg C	PSEG01	L31	FM782	08/26/2020	1010A

Date/Time 08/26/2020 12:05
 Received by: KS (wetchem)
 Relinquished by: KS (wetchem)

Date/Time 08/26/2020 16:16
 Received by: KS (wetchem)
 Relinquished by: KS (wetchem)