


LB110706

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.062	0.062	67	0.000	67				
0.2	0.200	0.245	0.045	718	0.0 %	718				
2.5	2.500	2.461	-0.039	8568	0.0 %	8568				
5.0	5.000	4.850	-0.150	17033	0.0 %	17033				
7.5	7.500	7.507	0.007	26449	0.0 %	26449				
10.0	10.000	10.075	0.075	35546	0.0 %	35546				

LB110706 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Aug 2020 09:18:18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	67	-	0.0000	7470A	PPB	27 Aug 2020 09:21:15	S	Std
0.2 - 1	50.2	718	-	0.2000	7470A	PPB	27 Aug 2020 09:23:32	S	Std
2.5 - 1	52.5	8568	-	2.5000	7470A	PPB	27 Aug 2020 09:25:49	S	Std
5.0 - 1	55	17033	-	5.0000	7470A	PPB	27 Aug 2020 09:28:06	S	Std
7.5 - 1	57.5	26449	-	7.5000	7470A	PPB	27 Aug 2020 09:30:23	S	Std
10.0 - 1	510	35546	-	10.0000	7470A	PPB	27 Aug 2020 09:37:58	S	Std
ICV86 - 1	ICV86	13863	3.9553	-7470A	PPB	27 Aug 2020 09:43:14	U	SMPL	
ICB86 - 1	ICB86	-181	-0.0084	-7470A	PPB	27 Aug 2020 09:45:28	U	SMPL	
CCV89 - 1	CCV89	17762	5.0557	-7470A	PPB	27 Aug 2020 09:47:44	U	SMPL	
CCB89 - 1	CCB89	-195	-0.0123	-7470A	PPB	27 Aug 2020 09:49:59	U	SMPL	
CRA - 1	CRA	653	0.2270	-7470A	PPB	27 Aug 2020 09:52:15	U	SMPL	
HighStd - 1	HighStd	35666	10.1087	-7470A	PPB	27 Aug 2020 09:54:31	U	SMPL	
ChkStd - 1	ChkStd	24733	7.0231	-7470A	PPB	27 Aug 2020 09:56:46	U	SMPL	
PB131217BL - 1	PBW	-311	-0.0450	-7470A	PPB	27 Aug 2020 09:59:05	U	SMPL	
PB131217BS - 1	LCSW	12117	3.4625	-7470A	PPB	27 Aug 2020 10:01:24	U	SMPL	
L3784-02 - 1	RO-10	-148	0.0010	-7470A	PPB	27 Aug 2020 10:03:40	U	SMPL	
L3784-02DUP - 1	RO-10DUP	4	0.0439	-7470A	PPB	27 Aug 2020 10:05:57	U	SMPL	
L3784-02MS - 1	RO-10MS	11512	3.2917	-7470A	PPB	27 Aug 2020 10:08:14	U	SMPL	
L3784-02MSD - 1	RO-10MSD	12512	3.5740	-7470A	PPB	27 Aug 2020 10:10:31	U	SMPL	
L3784-05 - 1	RO-11	-137	0.0041	-7470A	PPB	27 Aug 2020 10:12:47	U	SMPL	
CCV90 - 1	CCV90	17839	5.0774	-7470A	PPB	27 Aug 2020 10:15:03	U	SMPL	
CCB90 - 1	CCB90	-159	-0.0021	-7470A	PPB	27 Aug 2020 10:17:18	U	SMPL	
L3784-08 - 1	RO-23	-7	0.0408	-7470A	PPB	27 Aug 2020 10:19:35	U	SMPL	
L3784-11 - 1	RO-24	51	0.0571	-7470A	PPB	27 Aug 2020 10:21:50	U	SMPL	
L3784-14 - 1	RO-06	125	0.0780	-7470A	PPB	27 Aug 2020 10:24:06	U	SMPL	
L3784-17 - 1	RO-04	37	0.0532	-7470A	PPB	27 Aug 2020 10:26:22	U	SMPL	
L3785-02 - 1	CHRT-09	68	0.0619	-7470A	PPB	27 Aug 2020 10:28:38	U	SMPL	
L3785-05 - 1	CHRT-18	76	0.0642	-7470A	PPB	27 Aug 2020 10:30:54	U	SMPL	
PB131217TB - 1	PB131194TB	54	0.0580	-7470A	PPB	27 Aug 2020 10:33:11	U	SMPL	
L3784-02LX5 - 1		60	0.0597	-7470A	PPB	27 Aug 2020 10:35:28	U	SMPL	
L3784-02A - 1		13276	3.7896	-7470A	PPB	27 Aug 2020 10:37:45	U	SMPL	
CCV91 - 1	CCV91	17495	4.9803	-7470A	PPB	27 Aug 2020 10:40:01	U	SMPL	
CCB91 - 1	CCB91	-163	-0.0033	-7470A	PPB	27 Aug 2020 10:42:17	U	SMPL	