



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Alexander

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB110707

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.9	4.01	08/26/2020	16:10
2	CAL2	1	Water	NA	NA	22.9	7.01	08/26/2020	16:12
3	CAL3	1	Water	NA	NA	22.9	10.03	08/26/2020	16:14
4	ICV	1	Water	NA	NA	22.9	6.99	08/26/2020	16:16
5	CCV1	1	Water	NA	NA	23.1	2.01	08/26/2020	16:19
6	L3498-05	1	Solid	20.02	20	23.7	8.78	08/26/2020	16:22
7	L3498-05DUP	1	Solid	20.05	20	23.7	8.75	08/26/2020	16:24
8	CCV2	1	Water	NA	NA	23.1	12.01	08/26/2020	16:26

LB110767

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Corrosivity-solid-826-2

WorkList ID : 141817

Department : Wet-Chemistry

Date : 08-26-2020 15:39:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/26/2020	Solid	L3498-05	Corrosivity	Cool 4 deg C	PSEG05	L32	HR-06-082620	08/26/2020	9045D

Date/Time 8/26/2020 16:05
Received by: AMLiwet-chem
Relinquished by: CL (S07)

Date/Time 8/26/2020 16:30
Received by: CL (S07)
Relinquished by: AMLiwet-chem